

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP113023\
 Data File : PP062035.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Nov 2023 22:22
 Operator : YP\AJ
 Sample : 05575-23
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 60

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 01 00:16:04 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 28 04:28:00 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.452	3.724	26135833	33875985	19.853	18.826
2) SA Decachlor...	10.240	8.846	22652594	41549428	21.950	20.131
Target Compounds						
26) L6 AR-1254-1	6.485	5.654	6351771	11771417	113.217	108.110
27) L6 AR-1254-2	6.704	5.804	8179655	9245750	98.101	92.124
28) L6 AR-1254-3	7.071	6.215	5388931	9458657	65.139	60.545
29) L6 AR-1254-4	7.358	6.445	3148114	4280061	60.032	53.850
30) L6 AR-1254-5	7.778	6.869	2109696	4670192	33.119	33.092



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP113023\
Data File : PP062035.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Nov 2023 22:22
Operator : YP\AJ
Sample : 05575-23
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
60

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 01 00:16:04 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112723.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 28 04:28:00 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.452 min
Delta R.T.: 0.000 min
Response: 26135833
Conc: 19.85 ng/ml

Instrument :
ECD_P
ClientSampleId :
60




#1 Tetrachloro-m-xylene

R.T.: 3.724 min
Delta R.T.: -0.002 min
Response: 33875985
Conc: 18.83 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.240 min
Delta R.T.: -0.008 min
Response: 22652594
Conc: 21.95 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.846 min
Delta R.T.: -0.009 min
Response: 41549428
Conc: 20.13 ng/ml



#26 AR-1254-1


R.T.: 6.485 min
Delta R.T.: -0.003 min
Response: 6351771
Conc: 113.22 ng/ml

Instrument :
ECD_P
ClientSampleId :
60




#26 AR-1254-1

R.T.: 5.654 min
Delta R.T.: -0.003 min
Response: 11771417
Conc: 108.11 ng/ml




#27 AR-1254-2

R.T.: 6.704 min
Delta R.T.: -0.003 min
Response: 8179655
Conc: 98.10 ng/ml



#27 AR-1254-2


R.T.: 5.804 min
Delta R.T.: -0.003 min
Response: 9245750
Conc: 92.12 ng/ml



#28 AR-1254-3


R.T.: 7.071 min
Delta R.T.: -0.004 min
Response: 5388931
Conc: 65.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
60




#28 AR-1254-3

R.T.: 6.215 min
Delta R.T.: -0.003 min
Response: 9458657
Conc: 60.54 ng/ml




#29 AR-1254-4

R.T.: 7.358 min
Delta R.T.: -0.004 min
Response: 3148114
Conc: 60.03 ng/ml



#29 AR-1254-4


R.T.: 6.445 min
Delta R.T.: -0.004 min
Response: 4280061
Conc: 53.85 ng/ml



#30 AR-1254-5

R.T.: 7.778 min
Delta R.T.: -0.004 min
Response: 2109696
Conc: 33.12 ng/ml

Instrument :
ECD_P
ClientSampleId :
60



#30 AR-1254-5

R.T.: 6.869 min
Delta R.T.: -0.004 min
Response: 4670192
Conc: 33.09 ng/ml