

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP113023\
 Data File : PP062042.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Dec 2023 00:25
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 01 01:28:58 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 28 04:28:00 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.452	3.724	68586139	90415630	52.099	50.248
2) SA Decachlor...	10.241	8.847	50823595	96483898	49.247	46.747
Target Compounds						
26) L6 AR-1254-1	6.485	5.654	27433833	54448685	488.992	500.062
27) L6 AR-1254-2	6.705	5.803	43093034	49650942	516.825	494.717
28) L6 AR-1254-3	7.072	6.214	43045806	76201619	520.322	487.767
29) L6 AR-1254-4	7.359	6.446	28159391	40582698	536.976	510.592
30) L6 AR-1254-5	7.779	6.870	32917102	70250612	516.752	497.782



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP113023\
Data File : PP062042.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Dec 2023 00:25
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 01 01:28:58 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112723.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 28 04:28:00 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.452 min
Delta R.T.: -0.001 min
Response: 68586139
Conc: 52.10 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.724 min
Delta R.T.: -0.002 min
Response: 90415630
Conc: 50.25 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.241 min
Delta R.T.: -0.008 min
Response: 50823595
Conc: 49.25 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.847 min
Delta R.T.: -0.009 min
Response: 96483898
Conc: 46.75 ng/ml



#26 AR-1254-1


R.T.: 6.485 min
Delta R.T.: -0.003 min
Response: 27433833
Conc: 488.99 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.654 min
Delta R.T.: -0.003 min
Response: 54448685
Conc: 500.06 ng/ml




#27 AR-1254-2

R.T.: 6.705 min
Delta R.T.: -0.003 min
Response: 43093034
Conc: 516.82 ng/ml



#27 AR-1254-2


R.T.: 5.803 min
Delta R.T.: -0.003 min
Response: 49650942
Conc: 494.72 ng/ml



#28 AR-1254-3


R.T.: 7.072 min
Delta R.T.: -0.003 min
Response: 43045806
Conc: 520.32 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.214 min
Delta R.T.: -0.003 min
Response: 76201619
Conc: 487.77 ng/ml




#29 AR-1254-4

R.T.: 7.359 min
Delta R.T.: -0.003 min
Response: 28159391
Conc: 536.98 ng/ml



#29 AR-1254-4


R.T.: 6.446 min
Delta R.T.: -0.003 min
Response: 40582698
Conc: 510.59 ng/ml



#30 AR-1254-5

R.T.: 7.779 min
Delta R.T.: -0.003 min
Response: 32917102
Conc: 516.75 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.870 min
Delta R.T.: -0.003 min
Response: 70250612
Conc: 497.78 ng/ml