

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120121\
 Data File : PP041520.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 01 Dec 2021 21:46
 Operator : AJ\MA
 Sample : PB141097BSD
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 02 00:51:48 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP113021.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 30 23:30:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.745	3.750	482286	655959	21.388	21.595
2)	SA Decachlor...	10.638	9.007	478141	336711	22.206	21.786
Target Compounds							
3)	L1 AR-1016-1	6.060	4.997	341172	406497	426.392	437.547
4)	L1 AR-1016-2	6.084	5.015	518842	610579	435.464	430.112
5)	L1 AR-1016-3	6.149	5.206	319289	337975	428.474	420.956
6)	L1 AR-1016-4	6.256	5.259	258591	266911	425.968	423.586
7)	L1 AR-1016-5	6.569	5.487	252758	337191	431.329	453.300
8)	L2 AR-1221-1	4.995	4.006	56006	49979	213.078	142.416 #
9)	L2 AR-1221-2	5.094	4.105	44078	69587	235.200	252.777
10)	L2 AR-1221-3	5.181	4.192	185714	266701	297.816	307.312
11)	L3 AR-1232-1	5.181	4.192	185714	266701	359.389	360.042
12)	L3 AR-1232-2	5.765	5.015	264417	610579	1027.566	983.871
13)	L3 AR-1232-3	6.084	5.206	518842	337975	1048.459	1043.471
14)	L3 AR-1232-4	6.256	5.303	258591	327371	1053.328	1146.744
15)	L3 AR-1232-5	6.354	5.487	200117	337191	1243.068	1095.840
16)	L4 AR-1242-1	6.060	4.997	341172	406497	568.865	551.970
17)	L4 AR-1242-2	6.084	5.015	518842	610579	565.111	548.186
18)	L4 AR-1242-3	6.149	5.206	319289	337975	549.963	550.281
19)	L4 AR-1242-4	6.256	5.303	258591	327371	551.830	557.851
20)	L4 AR-1242-5	7.038	5.864	35775	255225	74.444	392.154 #
21)	L5 AR-1248-1	6.060	4.997	341172	406497	779.333	743.808
22)	L5 AR-1248-2	6.354	5.259	200117	266911	322.688	327.843
23)	L5 AR-1248-3	6.569	5.303	252758	327371	338.318	386.566
24)	L5 AR-1248-4	6.997	5.487	40828	337191	48.741	345.759 #
25)	L5 AR-1248-5	7.038	5.906	35775	41208	42.990	46.023
26)	L6 AR-1254-1	6.966	5.864	186816	255225	220.119	183.345
27)	L6 AR-1254-2	7.195	6.024	206616	235228	156.837	196.968 #
28)	L6 AR-1254-3	7.576	6.460	108747	401407	76.346	226.749 #
29)	L6 AR-1254-4	7.870	6.683	73448	44948	70.793	43.610 #
30)	L6 AR-1254-5	8.298	7.110	678396	692213	604.452	474.750
31)	L7 AR-1260-1	7.741	6.579	474207	543689	480.360	474.655
32)	L7 AR-1260-2	8.008	6.778	571589	611373	454.938	465.752
33)	L7 AR-1260-3	8.372	6.930	352854	598614	389.668	447.537
34)	L7 AR-1260-4	8.601	7.413	437185	403229	408.536	409.517
35)	L7 AR-1260-5	8.917	7.663	812038	852963	391.915	412.659
36)	L8 AR-1262-1	8.372	7.110	352854	692213	277.876	926.143 #
37)	L8 AR-1262-2	8.917	7.663	812038	852963	359.413	404.283
38)	L8 AR-1262-3	9.232f	7.947	569411	175521	524.620	196.956 #
39)	L8 AR-1262-4	9.284f	8.011	333927	620452	487.888	386.335
40)	L8 AR-1262-5	9.933	8.512	219522	186309	245.354	256.631
41)	L9 AR-1268-1	9.232f	7.947	569411	175521	210.381	73.185 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120121\
 Data File : PP041520.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 01 Dec 2021 21:46
 Operator : AJ\MA
 Sample : PB141097BSD
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 02 00:51:48 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP113021.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 30 23:30:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	9.284f	8.011	333927	620452	127.517	273.973 #
43) L9	AR-1268-3	9.517	8.216	15273	15080	6.891	8.030
44) L9	AR-1268-4	9.933	8.512	219522	186309	223.877	232.474
45) L9	AR-1268-5	10.324	8.783	74425	61563	10.296	12.035



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120121\
Data File : PP041520.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 01 Dec 2021 21:46
Operator : AJ\MA
Sample : PB141097BSD
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 02 00:51:48 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP113021.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 30 23:30:32 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.745 min
Delta R.T.: -0.002 min
Response: 482286
Conc: 21.39 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.750 min
Delta R.T.: -0.002 min
Response: 655959
Conc: 21.59 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.638 min
Delta R.T.: -0.006 min
Response: 478141
Conc: 22.21 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.007 min
Delta R.T.: -0.005 min
Response: 336711
Conc: 21.79 ng/ml



#3 AR-1016-1


R.T.: 6.060 min
Delta R.T.: -0.003 min
Response: 341172
Conc: 426.39 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.997 min
Delta R.T.: -0.002 min
Response: 406497
Conc: 437.55 ng/ml




#4 AR-1016-2

R.T.: 6.084 min
Delta R.T.: -0.003 min
Response: 518842
Conc: 435.46 ng/ml



#4 AR-1016-2


R.T.: 5.015 min
Delta R.T.: -0.002 min
Response: 610579
Conc: 430.11 ng/ml



#5 AR-1016-3


R.T.: 6.149 min
Delta R.T.: -0.002 min
Response: 319289
Conc: 428.47 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 5.206 min
Delta R.T.: -0.002 min
Response: 337975
Conc: 420.96 ng/ml




#6 AR-1016-4

R.T.: 6.256 min
Delta R.T.: -0.003 min
Response: 258591
Conc: 425.97 ng/ml



#6 AR-1016-4


R.T.: 5.259 min
Delta R.T.: -0.002 min
Response: 266911
Conc: 423.59 ng/ml



#7 AR-1016-5


R.T.: 6.569 min
Delta R.T.: -0.003 min
Response: 252758
Conc: 431.33 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.487 min
Delta R.T.: -0.002 min
Response: 337191
Conc: 453.30 ng/ml




#8 AR-1221-1

R.T.: 4.995 min
Delta R.T.: -0.002 min
Response: 56006
Conc: 213.08 ng/ml



#8 AR-1221-1


R.T.: 4.006 min
Delta R.T.: -0.003 min
Response: 49979
Conc: 142.42 ng/ml



#9 AR-1221-2


R.T.: 5.094 min
Delta R.T.: -0.002 min
Response: 44078
Conc: 235.20 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 4.105 min
Delta R.T.: -0.003 min
Response: 69587
Conc: 252.78 ng/ml




#10 AR-1221-3

R.T.: 5.181 min
Delta R.T.: -0.002 min
Response: 185714
Conc: 297.82 ng/ml



#10 AR-1221-3


R.T.: 4.192 min
Delta R.T.: -0.002 min
Response: 266701
Conc: 307.31 ng/ml



#11 AR-1232-1


R.T.: 5.181 min
Delta R.T.: -0.001 min
Response: 185714
Conc: 359.39 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 4.192 min
Delta R.T.: -0.003 min
Response: 266701
Conc: 360.04 ng/ml




#12 AR-1232-2

R.T.: 5.765 min
Delta R.T.: -0.003 min
Response: 264417
Conc: 1027.57 ng/ml



#12 AR-1232-2


R.T.: 5.015 min
Delta R.T.: -0.003 min
Response: 610579
Conc: 983.87 ng/ml



#13 AR-1232-3


R.T.: 6.084 min
Delta R.T.: -0.003 min
Response: 518842
Conc: 1048.46 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 5.206 min
Delta R.T.: -0.003 min
Response: 337975
Conc: 1043.47 ng/ml




#14 AR-1232-4

R.T.: 6.256 min
Delta R.T.: -0.003 min
Response: 258591
Conc: 1053.33 ng/ml



#14 AR-1232-4


R.T.: 5.303 min
Delta R.T.: -0.003 min
Response: 327371
Conc: 1146.74 ng/ml



#15 AR-1232-5


R.T.: 6.354 min
Delta R.T.: -0.002 min
Response: 200117
Conc: 1243.07 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.487 min
Delta R.T.: -0.003 min
Response: 337191
Conc: 1095.84 ng/ml




#16 AR-1242-1

R.T.: 6.060 min
Delta R.T.: -0.002 min
Response: 341172
Conc: 568.87 ng/ml



#16 AR-1242-1


R.T.: 4.997 min
Delta R.T.: -0.003 min
Response: 406497
Conc: 551.97 ng/ml



#17 AR-1242-2


R.T.: 6.084 min
Delta R.T.: -0.002 min
Response: 518842
Conc: 565.11 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 5.015 min
Delta R.T.: -0.002 min
Response: 610579
Conc: 548.19 ng/ml




#18 AR-1242-3

R.T.: 6.149 min
Delta R.T.: -0.003 min
Response: 319289
Conc: 549.96 ng/ml



#18 AR-1242-3


R.T.: 5.206 min
Delta R.T.: -0.003 min
Response: 337975
Conc: 550.28 ng/ml



#19 AR-1242-4


R.T.: 6.256 min
Delta R.T.: -0.002 min
Response: 258591
Conc: 551.83 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 5.303 min
Delta R.T.: -0.003 min
Response: 327371
Conc: 557.85 ng/ml




#20 AR-1242-5

R.T.: 7.038 min
Delta R.T.: -0.002 min
Response: 35775
Conc: 74.44 ng/ml



#20 AR-1242-5


R.T.: 5.864 min
Delta R.T.: -0.004 min
Response: 255225
Conc: 392.15 ng/ml



#21 AR-1248-1


R.T.: 6.060 min
Delta R.T.: -0.004 min
Response: 341172
Conc: 779.33 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.997 min
Delta R.T.: -0.004 min
Response: 406497
Conc: 743.81 ng/ml




#22 AR-1248-2

R.T.: 6.354 min
Delta R.T.: -0.003 min
Response: 200117
Conc: 322.69 ng/ml



#22 AR-1248-2


R.T.: 5.259 min
Delta R.T.: -0.004 min
Response: 266911
Conc: 327.84 ng/ml



#23 AR-1248-3


R.T.: 6.569 min
Delta R.T.: -0.004 min
Response: 252758
Conc: 338.32 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.303 min
Delta R.T.: -0.005 min
Response: 327371
Conc: 386.57 ng/ml




#24 AR-1248-4

R.T.: 6.997 min
Delta R.T.: -0.005 min
Response: 40828
Conc: 48.74 ng/ml



#24 AR-1248-4


R.T.: 5.487 min
Delta R.T.: -0.004 min
Response: 337191
Conc: 345.76 ng/ml



#25 AR-1248-5


R.T.: 7.038 min
Delta R.T.: -0.004 min
Response: 35775
Conc: 42.99 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.906 min
Delta R.T.: -0.005 min
Response: 41208
Conc: 46.02 ng/ml




#26 AR-1254-1

R.T.: 6.966 min
Delta R.T.: -0.004 min
Response: 186816
Conc: 220.12 ng/ml



#26 AR-1254-1


R.T.: 5.864 min
Delta R.T.: -0.005 min
Response: 255225
Conc: 183.35 ng/ml



#27 AR-1254-2


R.T.: 7.195 min
Delta R.T.: -0.006 min
Response: 206616
Conc: 156.84 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 6.024 min
Delta R.T.: -0.005 min
Response: 235228
Conc: 196.97 ng/ml




#28 AR-1254-3

R.T.: 7.576 min
Delta R.T.: -0.005 min
Response: 108747
Conc: 76.35 ng/ml



#28 AR-1254-3


R.T.: 6.460 min
Delta R.T.: 0.012 min
Response: 401407
Conc: 226.75 ng/ml



#29 AR-1254-4


R.T.: 7.870 min
Delta R.T.: -0.007 min
Response: 73448
Conc: 70.79 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.683 min
Delta R.T.: -0.006 min
Response: 44948
Conc: 43.61 ng/ml




#30 AR-1254-5

R.T.: 8.298 min
Delta R.T.: -0.005 min
Response: 678396
Conc: 604.45 ng/ml



#30 AR-1254-5


R.T.: 7.110 min
Delta R.T.: -0.006 min
Response: 692213
Conc: 474.75 ng/ml



#31 AR-1260-1


R.T.: 7.741 min
Delta R.T.: -0.004 min
Response: 474207
Conc: 480.36 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.579 min
Delta R.T.: -0.004 min
Response: 543689
Conc: 474.66 ng/ml




#32 AR-1260-2

R.T.: 8.008 min
Delta R.T.: -0.004 min
Response: 571589
Conc: 454.94 ng/ml



#32 AR-1260-2


R.T.: 6.778 min
Delta R.T.: -0.003 min
Response: 611373
Conc: 465.75 ng/ml



#33 AR-1260-3


R.T.: 8.372 min
Delta R.T.: -0.003 min
Response: 352854
Conc: 389.67 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.930 min
Delta R.T.: -0.004 min
Response: 598614
Conc: 447.54 ng/ml




#34 AR-1260-4

R.T.: 8.601 min
Delta R.T.: -0.005 min
Response: 437185
Conc: 408.54 ng/ml



#34 AR-1260-4


R.T.: 7.413 min
Delta R.T.: -0.005 min
Response: 403229
Conc: 409.52 ng/ml



#35 AR-1260-5


R.T.: 8.917 min
Delta R.T.: -0.004 min
Response: 812038
Conc: 391.91 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.663 min
Delta R.T.: -0.005 min
Response: 852963
Conc: 412.66 ng/ml




#36 AR-1262-1

R.T.: 8.372 min
Delta R.T.: -0.002 min
Response: 352854
Conc: 277.88 ng/ml



#36 AR-1262-1


R.T.: 7.110 min
Delta R.T.: -0.003 min
Response: 692213
Conc: 926.14 ng/ml



#37 AR-1262-2


R.T.: 8.917 min
Delta R.T.: -0.003 min
Response: 812038
Conc: 359.41 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 7.663 min
Delta R.T.: -0.003 min
Response: 852963
Conc: 404.28 ng/ml




#38 AR-1262-3

R.T.: 9.232 min
Delta R.T.: 0.018 min
Response: 569411
Conc: 524.62 ng/ml



#38 AR-1262-3


R.T.: 7.947 min
Delta R.T.: -0.003 min
Response: 175521
Conc: 196.96 ng/ml



#39 AR-1262-4


R.T.: 9.284 min
Delta R.T.: -0.019 min
Response: 333927
Conc: 487.89 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 8.011 min
Delta R.T.: -0.005 min
Response: 620452
Conc: 386.33 ng/ml




#40 AR-1262-5

R.T.: 9.933 min
Delta R.T.: -0.004 min
Response: 219522
Conc: 245.35 ng/ml



#40 AR-1262-5


R.T.: 8.512 min
Delta R.T.: -0.003 min
Response: 186309
Conc: 256.63 ng/ml



#41 AR-1268-1


R.T.: 9.232 min
Delta R.T.: 0.019 min
Response: 569411
Conc: 210.38 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.947 min
Delta R.T.: -0.003 min
Response: 175521
Conc: 73.19 ng/ml




#42 AR-1268-2

R.T.: 9.284 min
Delta R.T.: -0.022 min
Response: 333927
Conc: 127.52 ng/ml



#42 AR-1268-2


R.T.: 8.011 min
Delta R.T.: -0.006 min
Response: 620452
Conc: 273.97 ng/ml



#43 AR-1268-3


R.T.: 9.517 min
Delta R.T.: 0.000 min
Response: 15273
Conc: 6.89 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 8.216 min
Delta R.T.: -0.004 min
Response: 15080
Conc: 8.03 ng/ml




#44 AR-1268-4

R.T.: 9.933 min
Delta R.T.: -0.004 min
Response: 219522
Conc: 223.88 ng/ml



#44 AR-1268-4


R.T.: 8.512 min
Delta R.T.: -0.003 min
Response: 186309
Conc: 232.47 ng/ml



#45 AR-1268-5

R.T.: 10.324 min
Delta R.T.: -0.003 min
Response: 74425
Conc: 10.30 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.783 min
Delta R.T.: -0.003 min
Response: 61563
Conc: 12.04 ng/ml