

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120121\
 Data File : PP041497.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 01 Dec 2021 14:54
 Operator : AJ\MA
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 02 00:39:37 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP113021.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 30 23:30:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.745	3.749	1127653	1508222	50.008	49.652
2) SA Decachlor...	10.638	9.007	1066734	717497	49.541	46.424
Target Compounds						
3) L1 AR-1016-1	6.061	4.997	381974	456110	477.386	490.949
4) L1 AR-1016-2	6.084	5.015	595327	677674	499.658	477.375
5) L1 AR-1016-3	6.150	5.206	359881	389721	482.946	485.407
6) L1 AR-1016-4	6.257	5.260	289833	309306	477.432	490.867
7) L1 AR-1016-5	6.570	5.486	293129	360656	500.221	484.846
31) L7 AR-1260-1	7.741	6.580	526065	576242	532.890	503.075
32) L7 AR-1260-2	8.008	6.779	624383	638321	496.957	486.281
33) L7 AR-1260-3	8.371	6.932	466965	608313	515.684	454.788
34) L7 AR-1260-4	8.601	7.413	539746	474837	504.377	482.241
35) L7 AR-1260-5	8.916	7.664	1034086	1004268	499.082	485.860



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120121\
Data File : PP041497.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 01 Dec 2021 14:54
Operator : AJ\MA
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 02 00:39:37 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP113021.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 30 23:30:32 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.745 min
Delta R.T.: -0.002 min
Response: 1127653
Conc: 50.01 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.749 min
Delta R.T.: -0.002 min
Response: 1508222
Conc: 49.65 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.638 min
Delta R.T.: -0.006 min
Response: 1066734
Conc: 49.54 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.007 min
Delta R.T.: -0.005 min
Response: 717497
Conc: 46.42 ng/ml



#3 AR-1016-1


R.T.: 6.061 min
Delta R.T.: -0.002 min
Response: 381974
Conc: 477.39 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.997 min
Delta R.T.: -0.002 min
Response: 456110
Conc: 490.95 ng/ml




#4 AR-1016-2

R.T.: 6.084 min
Delta R.T.: -0.002 min
Response: 595327
Conc: 499.66 ng/ml



#4 AR-1016-2


R.T.: 5.015 min
Delta R.T.: -0.002 min
Response: 677674
Conc: 477.38 ng/ml



#5 AR-1016-3


R.T.: 6.150 min
Delta R.T.: -0.002 min
Response: 359881
Conc: 482.95 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 5.206 min
Delta R.T.: -0.002 min
Response: 389721
Conc: 485.41 ng/ml




#6 AR-1016-4

R.T.: 6.257 min
Delta R.T.: -0.002 min
Response: 289833
Conc: 477.43 ng/ml



#6 AR-1016-4


R.T.: 5.260 min
Delta R.T.: -0.002 min
Response: 309306
Conc: 490.87 ng/ml



#7 AR-1016-5


R.T.: 6.570 min
Delta R.T.: -0.002 min
Response: 293129
Conc: 500.22 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.486 min
Delta R.T.: -0.003 min
Response: 360656
Conc: 484.85 ng/ml




#31 AR-1260-1

R.T.: 7.741 min
Delta R.T.: -0.004 min
Response: 526065
Conc: 532.89 ng/ml



#31 AR-1260-1


R.T.: 6.580 min
Delta R.T.: -0.003 min
Response: 576242
Conc: 503.07 ng/ml



#32 AR-1260-2


R.T.: 8.008 min
Delta R.T.: -0.004 min
Response: 624383
Conc: 496.96 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.779 min
Delta R.T.: -0.003 min
Response: 638321
Conc: 486.28 ng/ml




#33 AR-1260-3

R.T.: 8.371 min
Delta R.T.: -0.004 min
Response: 466965
Conc: 515.68 ng/ml



#33 AR-1260-3


R.T.: 6.932 min
Delta R.T.: -0.003 min
Response: 608313
Conc: 454.79 ng/ml



#34 AR-1260-4


R.T.: 8.601 min
Delta R.T.: -0.005 min
Response: 539746
Conc: 504.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.413 min
Delta R.T.: -0.004 min
Response: 474837
Conc: 482.24 ng/ml



#35 AR-1260-5

R.T.: 8.916 min
Delta R.T.: -0.005 min
Response: 1034086
Conc: 499.08 ng/ml



#35 AR-1260-5

R.T.: 7.664 min
Delta R.T.: -0.004 min
Response: 1004268
Conc: 485.86 ng/ml