

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120121\
 Data File : PP041530.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 02 Dec 2021 1:00
 Operator : AJ\MA
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 02 01:05:17 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP113021.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 30 23:30:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.745	3.750	1187848	1679501	52.677	55.291
2) SA Decachlor...	10.638	9.007	1099617	754999	51.068	48.851
Target Compounds						
26) L6 AR-1254-1	6.967	5.864	430326	680926	507.038	489.155
27) L6 AR-1254-2	7.195	6.025	662871	586991	503.167	491.516
28) L6 AR-1254-3	7.576	6.443	701053	873186	492.178	493.250
29) L6 AR-1254-4	7.872	6.684	511040	512411	492.569	497.152
30) L6 AR-1254-5	8.299	7.112	549118	711959	489.265	488.292



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120121\
Data File : PP041530.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 02 Dec 2021 1:00
Operator : AJ\MA
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 02 01:05:17 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP113021.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 30 23:30:32 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.745 min
Delta R.T.: -0.002 min
Response: 1187848
Conc: 52.68 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.750 min
Delta R.T.: -0.002 min
Response: 1679501
Conc: 55.29 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.638 min
Delta R.T.: -0.006 min
Response: 1099617
Conc: 51.07 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.007 min
Delta R.T.: -0.006 min
Response: 754999
Conc: 48.85 ng/ml



#26 AR-1254-1


R.T.: 6.967 min
Delta R.T.: -0.003 min
Response: 430326
Conc: 507.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.864 min
Delta R.T.: -0.005 min
Response: 680926
Conc: 489.15 ng/ml




#27 AR-1254-2

R.T.: 7.195 min
Delta R.T.: -0.005 min
Response: 662871
Conc: 503.17 ng/ml



#27 AR-1254-2


R.T.: 6.025 min
Delta R.T.: -0.004 min
Response: 586991
Conc: 491.52 ng/ml



#28 AR-1254-3


R.T.: 7.576 min
Delta R.T.: -0.005 min
Response: 701053
Conc: 492.18 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.443 min
Delta R.T.: -0.005 min
Response: 873186
Conc: 493.25 ng/ml




#29 AR-1254-4

R.T.: 7.872 min
Delta R.T.: -0.005 min
Response: 511040
Conc: 492.57 ng/ml



#29 AR-1254-4


R.T.: 6.684 min
Delta R.T.: -0.005 min
Response: 512411
Conc: 497.15 ng/ml



#30 AR-1254-5

R.T.: 8.299 min
Delta R.T.: -0.005 min
Response: 549118
Conc: 489.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 7.112 min
Delta R.T.: -0.005 min
Response: 711959
Conc: 488.29 ng/ml