

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120122\
 Data File : PP053604.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Dec 2022 15:06
 Operator : YP\AJ
 Sample : N5823-01
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 OE-2-113022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 01 20:43:12 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 09 05:50:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.398	3.644	46237182	42883092	24.053	28.508
2) SA Decachlor...	10.197	8.720	35171292	41175147	25.938	21.968

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120122\
Data File : PP053604.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Dec 2022 15:06
Operator : YP\AJ
Sample : N5823-01
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
OE-2-113022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 01 20:43:12 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110822.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 09 05:50:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.398 min
Delta R.T.: -0.005 min
Response: 46237182
Conc: 24.05 ng/ml

Instrument :
ECD_P
ClientSampleId :
OE-2-113022




#1 Tetrachloro-m-xylene

R.T.: 3.644 min
Delta R.T.: -0.005 min
Response: 42883092
Conc: 28.51 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.197 min
Delta R.T.: -0.005 min
Response: 35171292
Conc: 25.94 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.720 min
Delta R.T.: -0.009 min
Response: 41175147
Conc: 21.97 ng/ml