

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120123\
 Data File : PP062079.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Dec 2023 16:56
 Operator : YP\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 01 17:08:40 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 28 04:28:00 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.453	3.724	27861429	35628885	21.164	19.801
2) SA Decachlor...	10.242	8.844	22819435	40784764	22.112	19.761
Target Compounds						
9) L2 AR-1221-2	4.761	4.031	510102	516229	38.249	27.140 #
32) L7 AR-1260-2	0.000	6.517f	0	230583	N.D.	1.692 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120123\
Data File : PP062079.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Dec 2023 16:56
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 01 17:08:40 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112723.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 28 04:28:00 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.453 min
Delta R.T.: 0.000 min
Response: 27861429
Conc: 21.16 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.724 min
Delta R.T.: -0.002 min
Response: 35628885
Conc: 19.80 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.242 min
Delta R.T.: -0.006 min
Response: 22819435
Conc: 22.11 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.844 min
Delta R.T.: -0.011 min
Response: 40784764
Conc: 19.76 ng/ml



#9 AR-1221-2


R.T.: 4.761 min
Delta R.T.: 0.011 min
Response: 510102
Conc: 38.25 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 4.031 min
Delta R.T.: -0.001 min
Response: 516229
Conc: 27.14 ng/ml



#32 AR-1260-2

R.T.: 0.000 min
Exp R.T. : 7.499 min
Response: 0
Conc: N.D.



#32 AR-1260-2

R.T.: 6.517 min
Delta R.T.: -0.024 min
Response: 230583
Conc: 1.69 ng/ml