

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120125\
 Data File : PP076654.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Dec 2025 12:20
 Operator : YP\AJ
 Sample : Q3530-07
 Misc : LOD-MDL-1660-WATER-01-QT4-2025
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 LOD-MDL-WATER-01-QT4-2025

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 12/02/2025
 Supervised By :Sohil Jodhani 12/03/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 02 00:53:10 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Nov 21 22:52:01 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.617	3.774	40398653	52972764	22.010	25.432
2) SA Decachlor...	10.365	8.768	33215248	43330722	22.163	24.557
Target Compounds						
3) L1 AR-1016-1	5.767	4.853	1777890	2536610	26.333m	33.745 #
4) L1 AR-1016-2	5.789	4.870	3057775	3554897	30.303m	31.971
5) L1 AR-1016-3	5.851	5.044	2044629	1948285	31.458m	33.612m
6) L1 AR-1016-4	5.949	5.092	1693601	1414365	31.630m	30.253m
7) L1 AR-1016-5	6.239	5.298	1485581	2189323	28.808m	35.346m
31) L7 AR-1260-1	7.357	6.325	2766949	3819615	31.857m	35.010m
32) L7 AR-1260-2	7.610	6.512	3423313	4805666	33.122m	36.138m
33) L7 AR-1260-3	7.967	6.666	2483038	4524846	30.889m	36.522
34) L7 AR-1260-4	8.193	7.135	3474927	3574558	37.097m	34.411
35) L7 AR-1260-5	8.517	7.376	5565508	8099256	32.735	32.774

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120125\
Data File : PP076654.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Dec 2025 12:20
Operator : YP\AJ
Sample : Q3530-07
Misc : LOD-MDL-1660-WATER-01-QT4-2025
ALS Vial : 6 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
LOD-MDL-WATER-01-QT4-2025


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 12/02/2025
Supervised By :Sohil Jodhani 12/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 02 00:53:10 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Nov 21 22:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.617 min
Delta R.T.: -0.001 min
Response: 40398653
Conc: 22.01 ng/ml

Instrument :
ECD_P
ClientSampleId :
LOD-MDL-WATER-01-QT4-2025


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 12/02/2025
Supervised By :Sohil Jodhani 12/03/2025




#1 Tetrachloro-m-xylene

R.T.: 3.774 min
Delta R.T.: 0.000 min
Response: 52972764
Conc: 25.43 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.365 min
Delta R.T.: -0.005 min
Response: 33215248
Conc: 22.16 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.768 min
Delta R.T.: -0.004 min
Response: 43330722
Conc: 24.56 ng/ml




#3 AR-1016-1

R.T.: 5.767 min
Delta R.T.: -0.004 min
Response: 1777890
Conc: 26.33 ng/ml

Instrument :
ECD_P
ClientSampleId :
LOD-MDL-WATER-01-QT4-2025


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 12/02/2025
Supervised By :Sohil Jodhani 12/03/2025




#3 AR-1016-1

R.T.: 4.853 min
Delta R.T.: 0.000 min
Response: 2536610
Conc: 33.74 ng/ml




#4 AR-1016-2

R.T.: 5.789 min
Delta R.T.: -0.003 min
Response: 3057775
Conc: 30.30 ng/ml m



#4 AR-1016-2

R.T.: 4.870 min
Delta R.T.: -0.001 min
Response: 3554897
Conc: 31.97 ng/ml




#5 AR-1016-3

R.T.: 5.851 min
Delta R.T.: -0.003 min
Response: 2044629
Conc: 31.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
LOD-MDL-WATER-01-QT4-2025


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 12/02/2025
Supervised By :Sohil Jodhani 12/03/2025




#5 AR-1016-3

R.T.: 5.044 min
Delta R.T.: -0.003 min
Response: 1948285
Conc: 33.61 ng/ml m




#6 AR-1016-4

R.T.: 5.949 min
Delta R.T.: -0.003 min
Response: 1693601
Conc: 31.63 ng/ml m



#6 AR-1016-4

R.T.: 5.092 min
Delta R.T.: 0.004 min
Response: 1414365
Conc: 30.25 ng/ml m




#7 AR-1016-5

R.T.: 6.239 min
Delta R.T.: -0.005 min
Response: 1485581
Conc: 28.81 ng/ml

Instrument :
ECD_P
ClientSampleId :
LOD-MDL-WATER-01-QT4-2025


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 12/02/2025
Supervised By :Sohil Jodhani 12/03/2025




#7 AR-1016-5

R.T.: 5.298 min
Delta R.T.: -0.003 min
Response: 2189323
Conc: 35.35 ng/ml m




#31 AR-1260-1

R.T.: 7.357 min
Delta R.T.: -0.004 min
Response: 2766949
Conc: 31.86 ng/ml m



#31 AR-1260-1

R.T.: 6.325 min
Delta R.T.: -0.003 min
Response: 3819615
Conc: 35.01 ng/ml m




#32 AR-1260-2

R.T.: 7.610 min
Delta R.T.: -0.004 min
Response: 3423313
Conc: 33.12 ng/ml

Instrument :
ECD_P
ClientSampleId :
LOD-MDL-WATER-01-QT4-2025


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 12/02/2025
Supervised By :Sohil Jodhani 12/03/2025




#32 AR-1260-2

R.T.: 6.512 min
Delta R.T.: -0.004 min
Response: 4805666
Conc: 36.14 ng/ml m




#33 AR-1260-3

R.T.: 7.967 min
Delta R.T.: -0.005 min
Response: 2483038
Conc: 30.89 ng/ml m



#33 AR-1260-3

R.T.: 6.666 min
Delta R.T.: -0.002 min
Response: 4524846
Conc: 36.52 ng/ml




#34 AR-1260-4

R.T.: 8.193 min
Delta R.T.: -0.005 min
Response: 3474927
Conc: 37.10 ng/ml

Instrument :
ECD_P
ClientSampleId :
LOD-MDL-WATER-01-QT4-2025


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 12/02/2025
Supervised By :Sohil Jodhani 12/03/2025




#34 AR-1260-4

R.T.: 7.135 min
Delta R.T.: -0.002 min
Response: 3574558
Conc: 34.41 ng/ml



#35 AR-1260-5

R.T.: 8.517 min
Delta R.T.: -0.004 min
Response: 5565508
Conc: 32.74 ng/ml



#35 AR-1260-5

R.T.: 7.376 min
Delta R.T.: -0.002 min
Response: 8099256
Conc: 32.77 ng/ml