

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120220\
 Data File : PP031779.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 02 Dec 2020 10:44
 Operator : DD\AJ
 Sample : L4905-01 2X
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 COMP-00-01-02

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 02 13:04:33 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112220.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 24 12:23:58 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.786	4.061	382876	283434	7.907	5.915 #
2)	SA Decachlor...	10.662	9.364	262295	260535	7.723	7.036
Target Compounds							
16)	L4 AR-1242-1	6.093	5.316	5769341	5360497	4821.074	4695.564
17)	L4 AR-1242-2	6.117	5.336	8461920	7863274	4755.799	4675.702
18)	L4 AR-1242-3	6.182	5.527	5659786	4626337	5083.203	5066.494
19)	L4 AR-1242-4	6.288	5.624	5016199	3771017	5427.608	4489.266
20)	L4 AR-1242-5	7.067	6.187	4710248	5744542	5141.444	5567.857
31)	L7 AR-1260-1	7.773	6.903	1493055	1510487	853.153	764.382
32)	L7 AR-1260-2	8.038	7.099	1607219	1552881	782.200	637.669
33)	L7 AR-1260-3	8.404	7.254	1176033	1535320	721.035	672.313
34)	L7 AR-1260-4	8.632	7.736	1458019	1326737	791.822	698.239
35)	L7 AR-1260-5	8.946	7.984	2761036	3075876	748.814	677.812



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120220\
Data File : PP031779.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 02 Dec 2020 10:44
Operator : DD\AJ
Sample : L4905-01 2X
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :
COMP-00-01-02

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 02 13:04:33 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112220.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 24 12:23:58 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.786 min
Delta R.T.: -0.004 min
Response: 382876
Conc: 7.91 ng/ml

Instrument :
ECD_P
ClientSampleId :
COMP-00-01-02




#1 Tetrachloro-m-xylene

R.T.: 4.061 min
Delta R.T.: 0.000 min
Response: 283434
Conc: 5.91 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.662 min
Delta R.T.: -0.003 min
Response: 262295
Conc: 7.72 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.364 min
Delta R.T.: -0.006 min
Response: 260535
Conc: 7.04 ng/ml



#16 AR-1242-1


R.T.: 6.093 min
Delta R.T.: -0.002 min
Response: 5769341
Conc: 4821.07 ng/ml

Instrument :
ECD_P
ClientSampleId :
COMP-00-01-02




#16 AR-1242-1

R.T.: 5.316 min
Delta R.T.: -0.002 min
Response: 5360497
Conc: 4695.56 ng/ml




#17 AR-1242-2

R.T.: 6.117 min
Delta R.T.: -0.001 min
Response: 8461920
Conc: 4755.80 ng/ml



#17 AR-1242-2


R.T.: 5.336 min
Delta R.T.: -0.003 min
Response: 7863274
Conc: 4675.70 ng/ml



#18 AR-1242-3


R.T.: 6.182 min
Delta R.T.: 0.000 min
Response: 5659786
Conc: 5083.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
COMP-00-01-02




#18 AR-1242-3

R.T.: 5.527 min
Delta R.T.: -0.003 min
Response: 4626337
Conc: 5066.49 ng/ml




#19 AR-1242-4

R.T.: 6.288 min
Delta R.T.: -0.001 min
Response: 5016199
Conc: 5427.61 ng/ml



#19 AR-1242-4


R.T.: 5.624 min
Delta R.T.: -0.002 min
Response: 3771017
Conc: 4489.27 ng/ml



#20 AR-1242-5


R.T.: 7.067 min
Delta R.T.: 0.000 min
Response: 4710248
Conc: 5141.44 ng/ml

Instrument :
ECD_P
ClientSampleId :
COMP-00-01-02




#20 AR-1242-5

R.T.: 6.187 min
Delta R.T.: -0.003 min
Response: 5744542
Conc: 5567.86 ng/ml




#31 AR-1260-1

R.T.: 7.773 min
Delta R.T.: -0.002 min
Response: 1493055
Conc: 853.15 ng/ml



#31 AR-1260-1


R.T.: 6.903 min
Delta R.T.: -0.004 min
Response: 1510487
Conc: 764.38 ng/ml



#32 AR-1260-2


R.T.: 8.038 min
Delta R.T.: -0.002 min
Response: 1607219
Conc: 782.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
COMP-00-01-02




#32 AR-1260-2

R.T.: 7.099 min
Delta R.T.: -0.004 min
Response: 1552881
Conc: 637.67 ng/ml




#33 AR-1260-3

R.T.: 8.404 min
Delta R.T.: -0.003 min
Response: 1176033
Conc: 721.04 ng/ml



#33 AR-1260-3


R.T.: 7.254 min
Delta R.T.: -0.004 min
Response: 1535320
Conc: 672.31 ng/ml



#34 AR-1260-4


R.T.: 8.632 min
Delta R.T.: -0.002 min
Response: 1458019
Conc: 791.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
COMP-00-01-02




#34 AR-1260-4

R.T.: 7.736 min
Delta R.T.: -0.005 min
Response: 1326737
Conc: 698.24 ng/ml



#35 AR-1260-5

R.T.: 8.946 min
Delta R.T.: -0.002 min
Response: 2761036
Conc: 748.81 ng/ml



#35 AR-1260-5

R.T.: 7.984 min
Delta R.T.: -0.004 min
Response: 3075876
Conc: 677.81 ng/ml