

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120325\
 Data File : PP076772.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Dec 2025 21:08
 Operator : YP\AJ
 Sample : Q3754-01 10X
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 L73A

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 12/04/2025
 Supervised By :mohammad ahmed 12/05/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 04 04:07:22 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Nov 21 22:52:01 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.614	3.771	2807031	2669511	1.529m	1.282
2) SA Decachlor...	10.364	8.763	3625788	2055334	2.419	1.165m#

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120325\
Data File : PP076772.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Dec 2025 21:08
Operator : YP\AJ
Sample : Q3754-01 10X
Misc :
ALS Vial : 25 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
L73A


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 12/04/2025
Supervised By :mohammad ahmed 12/05/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 04 04:07:22 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Nov 21 22:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.614 min
Delta R.T.: -0.004 min
Response: 2807031
Conc: 1.53 ng/ml

Instrument :
ECD_P
ClientSampleId :
L73A


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 12/04/2025
Supervised By :mohammad ahmed 12/05/2025




#1 Tetrachloro-m-xylene

R.T.: 3.771 min
Delta R.T.: -0.004 min
Response: 2669511
Conc: 1.28 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.364 min
Delta R.T.: -0.006 min
Response: 3625788
Conc: 2.42 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.763 min
Delta R.T.: -0.008 min
Response: 2055334
Conc: 1.16 ng/ml m