

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120418\
 Data File : PP028269.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 04 Dec 2018 17:40
 Operator : EI\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 05 14:07:38 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110918.M
 Quant Title : 8080.M
 QLast Update : Fri Nov 09 17:06:51 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	12.094	12.206	194.7E6	332.3E6	748.082	782.924
Target Compounds							
1) T	Dalapon	3.342	3.011	411.0E6	382.7E6	654.814	654.492
2) T	3,5-DICHL...	10.507	10.297	349.0E6	375.6E6	650.913	726.027
3) T	4-Nitroph...	11.670	11.351	315.4E6	156.2E6	599.295	686.848
5) T	DICAMBA	12.415	12.517	1153.7E6	1232.1E6	697.283	733.659
6) T	MCPD	12.758	12.708	29828684	61247253	64.711	73.778
7) T	MCPA	13.001	13.097	49073910	94578936	67.709	73.165
8) T	DICHLORPROP	13.613	13.632	230.0E6	301.4E6	537.594	731.677 #
9) T	2,4-D	13.964	14.119	271.4E6	357.5E6	706.913	739.418
10) T	Pentachlo...	14.383	14.731	5889.5E6	4740.9E6	703.207	735.407
11) T	2,4,5-TP ...	15.245	15.301	2112.5E6	1819.7E6	701.652	741.238
12) T	2,4,5-T	15.635	15.942	1919.7E6	1579.2E6	692.196	726.642
13) T	2,4-DB	16.400	16.545	159.1E6	194.8E6	700.983	736.953
14) T	DINOSEB	17.888	16.936	1540.4E6	1277.9E6	678.749	710.847
15) T	Picloram	17.652	18.316	3090.0E6	2592.1E6	652.694	730.771
16) T	DCPA	18.251	18.196	3134.9E6	2355.6E6	693.345	732.846

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120418\
Data File : PP028269.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 04 Dec 2018 17:40
Operator : EI\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 05 14:07:38 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110918.M
Quant Title : 8080.M
QLast Update : Fri Nov 09 17:06:51 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

