

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120423\
 Data File : PP062121.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Dec 2023 13:54
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 04 14:10:01 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 28 04:28:00 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.451	3.722	75873497	91277624	57.635	50.727
2) SA Decachlor...	10.243	8.847	55340150	89752106	53.624	43.486
Target Compounds						
21) L5 AR-1248-1	5.627	4.835	16374976	19891072	535.901	498.235
22) L5 AR-1248-2	5.902	5.077	24411439	30512216	498.992	483.075
23) L5 AR-1248-3	6.107	5.120	26081193	31985641	501.323	487.684
24) L5 AR-1248-4	6.513	5.295	26402054	37127865	520.893	482.329
25) L5 AR-1248-5	6.551	5.653	26163062	49640915	506.296	496.740



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120423\
Data File : PP062121.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Dec 2023 13:54
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 04 14:10:01 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112723.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 28 04:28:00 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.451 min
Delta R.T.: -0.002 min
Response: 75873497
Conc: 57.63 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.722 min
Delta R.T.: -0.004 min
Response: 91277624
Conc: 50.73 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.243 min
Delta R.T.: -0.005 min
Response: 55340150
Conc: 53.62 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.847 min
Delta R.T.: -0.008 min
Response: 89752106
Conc: 43.49 ng/ml



#21 AR-1248-1


R.T.: 5.627 min
Delta R.T.: -0.002 min
Response: 16374976
Conc: 535.90 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.835 min
Delta R.T.: -0.004 min
Response: 19891072
Conc: 498.23 ng/ml




#22 AR-1248-2

R.T.: 5.902 min
Delta R.T.: -0.001 min
Response: 24411439
Conc: 498.99 ng/ml



#22 AR-1248-2


R.T.: 5.077 min
Delta R.T.: -0.003 min
Response: 30512216
Conc: 483.08 ng/ml



#23 AR-1248-3


R.T.: 6.107 min
Delta R.T.: -0.002 min
Response: 26081193
Conc: 501.32 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.120 min
Delta R.T.: -0.003 min
Response: 31985641
Conc: 487.68 ng/ml




#24 AR-1248-4

R.T.: 6.513 min
Delta R.T.: -0.002 min
Response: 26402054
Conc: 520.89 ng/ml



#24 AR-1248-4


R.T.: 5.295 min
Delta R.T.: -0.003 min
Response: 37127865
Conc: 482.33 ng/ml



#25 AR-1248-5

R.T.: 6.551 min
Delta R.T.: -0.002 min
Response: 26163062
Conc: 506.30 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.653 min
Delta R.T.: -0.004 min
Response: 49640915
Conc: 496.74 ng/ml