

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120423\
 Data File : PP062131.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Dec 2023 17:00
 Operator : YP\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 I.BLK

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/05/2023
 Supervised By :Ankita Jodhani 12/05/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 04 17:16:27 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 28 04:28:00 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.451	3.722	26997011	33302771	20.507m	18.508
2) SA Decachlor...	10.244	8.847	22220321	38834602	21.531	18.816

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120423\
Data File : PP062131.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Dec 2023 17:00
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
I.BLK


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/05/2023
Supervised By :Ankita Jodhani 12/05/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 04 17:16:27 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112723.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 28 04:28:00 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





#1 Tetrachloro-m-xylene

R.T.: 4.451 min
Delta R.T.: -0.002 min
Response: 26997011
Conc: 20.51 ng/ml

Instrument :


ECD_P

ClientSampleId :

I.BLK


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/05/2023
Supervised By :Ankita Jodhani 12/05/2023




#1 Tetrachloro-m-xylene

R.T.: 3.722 min
Delta R.T.: -0.004 min
Response: 33302771
Conc: 18.51 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.244 min
Delta R.T.: -0.004 min
Response: 22220321
Conc: 21.53 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.847 min
Delta R.T.: -0.009 min
Response: 38834602
Conc: 18.82 ng/ml