

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120523\
 Data File : PP062133.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Dec 2023 09:07
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 06 00:18:43 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 28 04:28:00 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.452	3.721	73375770	88520414	55.737	49.195
2) SA Decachlor...	10.245	8.845	52223975	96265631	50.604	46.642
Target Compounds						
3) L1 AR-1016-1	5.627	4.835	26976631	33116379	544.409	499.901
4) L1 AR-1016-2	5.649	4.854	39605150	47004422	530.912	508.801
5) L1 AR-1016-3	5.712	5.035	26692790	25641271	542.392	505.451
6) L1 AR-1016-4	5.809	5.077	21882132	21755100	575.712	477.485
7) L1 AR-1016-5	6.108	5.294	20012848	28742026	520.403	497.172
31) L7 AR-1260-1	7.239	6.345	35166844	56021633	514.819	481.374
32) L7 AR-1260-2	7.497	6.535	39069016	66803812	518.699	490.293
33) L7 AR-1260-3	7.858	6.691	26060368	62962009	503.035	481.206
34) L7 AR-1260-4	8.084	7.170	30926810	47249863	506.772	479.122
35) L7 AR-1260-5	8.408	7.413	52581819	107.9E6	524.356	504.128



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120523\
Data File : PP062133.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Dec 2023 09:07
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 06 00:18:43 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112723.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 28 04:28:00 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.452 min
Delta R.T.: -0.001 min
Response: 73375770
Conc: 55.74 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.721 min
Delta R.T.: -0.005 min
Response: 88520414
Conc: 49.19 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.245 min
Delta R.T.: -0.003 min
Response: 52223975
Conc: 50.60 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.845 min
Delta R.T.: -0.010 min
Response: 96265631
Conc: 46.64 ng/ml



#3 AR-1016-1


R.T.: 5.627 min
Delta R.T.: -0.002 min
Response: 26976631
Conc: 544.41 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.835 min
Delta R.T.: -0.005 min
Response: 33116379
Conc: 499.90 ng/ml




#4 AR-1016-2

R.T.: 5.649 min
Delta R.T.: -0.003 min
Response: 39605150
Conc: 530.91 ng/ml



#4 AR-1016-2


R.T.: 4.854 min
Delta R.T.: -0.005 min
Response: 47004422
Conc: 508.80 ng/ml



#5 AR-1016-3


R.T.: 5.712 min
Delta R.T.: -0.002 min
Response: 26692790
Conc: 542.39 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 5.035 min
Delta R.T.: -0.005 min
Response: 25641271
Conc: 505.45 ng/ml




#6 AR-1016-4

R.T.: 5.809 min
Delta R.T.: -0.004 min
Response: 21882132
Conc: 575.71 ng/ml



#6 AR-1016-4


R.T.: 5.077 min
Delta R.T.: -0.005 min
Response: 21755100
Conc: 477.49 ng/ml



#7 AR-1016-5


R.T.: 6.108 min
Delta R.T.: -0.001 min
Response: 20012848
Conc: 520.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.294 min
Delta R.T.: -0.005 min
Response: 28742026
Conc: 497.17 ng/ml




#31 AR-1260-1

R.T.: 7.239 min
Delta R.T.: -0.002 min
Response: 35166844
Conc: 514.82 ng/ml



#31 AR-1260-1


R.T.: 6.345 min
Delta R.T.: -0.006 min
Response: 56021633
Conc: 481.37 ng/ml



#32 AR-1260-2


R.T.: 7.497 min
Delta R.T.: -0.002 min
Response: 39069016
Conc: 518.70 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.535 min
Delta R.T.: -0.006 min
Response: 66803812
Conc: 490.29 ng/ml




#33 AR-1260-3

R.T.: 7.858 min
Delta R.T.: -0.003 min
Response: 26060368
Conc: 503.04 ng/ml



#33 AR-1260-3


R.T.: 6.691 min
Delta R.T.: -0.006 min
Response: 62962009
Conc: 481.21 ng/ml



#34 AR-1260-4


R.T.: 8.084 min
Delta R.T.: -0.003 min
Response: 30926810
Conc: 506.77 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.170 min
Delta R.T.: -0.006 min
Response: 47249863
Conc: 479.12 ng/ml



#35 AR-1260-5

R.T.: 8.408 min
Delta R.T.: -0.003 min
Response: 52581819
Conc: 524.36 ng/ml



#35 AR-1260-5

R.T.: 7.413 min
Delta R.T.: -0.006 min
Response: 107947330
Conc: 504.13 ng/ml