

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120525\
 Data File : PP076832.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Dec 2025 14:24
 Operator : YP\AJ
 Sample : Q3786-01
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 CAMDEN GAS PIPE#1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 05 15:04:00 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Nov 21 22:52:01 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.616	3.771	31066773	35993330	16.925	17.280
2) SA Decachlor...	10.365	8.763	24417012	26015153	16.292	14.744

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120525\
Data File : PP076832.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Dec 2025 14:24
Operator : YP\AJ
Sample : Q3786-01
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
CAMDEN GAS PIPE#1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 05 15:04:00 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Nov 21 22:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.616 min
Delta R.T.: -0.002 min
Response: 31066773
Conc: 16.93 ng/ml

Instrument :
ECD_P
ClientSampleId :
CAMDEN GAS PIPE#1




#1 Tetrachloro-m-xylene

R.T.: 3.771 min
Delta R.T.: -0.004 min
Response: 35993330
Conc: 17.28 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.365 min
Delta R.T.: -0.005 min
Response: 24417012
Conc: 16.29 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.763 min
Delta R.T.: -0.008 min
Response: 26015153
Conc: 14.74 ng/ml