

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120621\
 Data File : PP041648.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 06 Dec 2021 15:23
 Operator : AJ\MA
 Sample : M4904-09 10X
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 KMA7700-1-1

Manual Integrations
 APPROVED

Reviewed By :Ankita Jodhani 12/07/2021
 Supervised By :mohammad ahmed 12/07/2021

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 06 23:55:18 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP113021.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 30 23:30:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.746	3.749	38046	57987	1.687	1.909
2) SA Decachlor...	10.643	9.008	45933	36908	2.133	2.388m
Target Compounds						
26) L6 AR-1254-1	6.968	5.865	18261	29858	21.516	21.449m
27) L6 AR-1254-2	7.198	6.025	26620	28954	20.207	24.245m
28) L6 AR-1254-3	7.578	6.443	29680	36665	20.837	20.711m
29) L6 AR-1254-4	7.875	6.684	22287	22538	21.481	21.867m
30) L6 AR-1254-5	8.301	7.112	29229	36299	26.043	24.895

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120621\
Data File : PP041648.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 06 Dec 2021 15:23
Operator : AJ\MA
Sample : M4904-09 10X
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
KMA7700-1-1

Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 12/07/2021
Supervised By :mohammad ahmed 12/07/2021

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 06 23:55:18 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP113021.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 30 23:30:32 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

