

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120623\
 Data File : PP062203.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Dec 2023 11:06
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/07/2023
 Supervised By :Ankita Jodhani 12/07/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 06 11:36:02 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 28 04:28:00 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.449	3.720	75256353	88678313	57.166	49.282
2)	SA Decachlor...	10.239	8.842	51709294	98462354	50.105m	47.706
Target Compounds							
3)	L1 AR-1016-1	5.625	4.833	26112108	32442513	526.962	489.729
4)	L1 AR-1016-2	5.647	4.853	38407883	46052036	514.862	498.492
5)	L1 AR-1016-3	5.709	5.033	24921312	25090840	506.396	494.600
6)	L1 AR-1016-4	5.808	5.075	19774236	21724482	520.254	476.813
7)	L1 AR-1016-5	6.105	5.292	19455221	27532449	505.903	476.249
31)	L7 AR-1260-1	7.236	6.343	33766621	55413266	494.321	476.146
32)	L7 AR-1260-2	7.494	6.533	37711915	65389752	500.681	479.915
33)	L7 AR-1260-3	7.855	6.689	25752312	61893866	497.089	473.042
34)	L7 AR-1260-4	8.081	7.167	30336952	47025583	497.106	476.848
35)	L7 AR-1260-5	8.404	7.411	50927772	105.1E6	507.861	490.950

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120623\
Data File : PP062203.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Dec 2023 11:06
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/07/2023
Supervised By :Ankita Jodhani 12/07/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 06 11:36:02 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112723.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 28 04:28:00 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

