

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120623\
 Data File : PP062219.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Dec 2023 16:26
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 07 00:22:34 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 28 04:28:00 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.449	3.720	72855246	86588727	55.342	48.121
2) SA Decachlor...	10.242	8.843	49225074	95669132	47.698	46.353
Target Compounds						
16) L4 AR-1242-1	5.625	4.833	19415926	24375099	488.540	460.031
17) L4 AR-1242-2	5.648	4.852	28532635	34295449	479.912	471.008
18) L4 AR-1242-3	5.710	5.033	18839253	18971009	481.789	468.132
19) L4 AR-1242-4	5.809	5.118	14823415	20851477	488.800	461.990
20) L4 AR-1242-5	6.550	5.650	14912982	25115209	497.392	479.938



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120623\
Data File : PP062219.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Dec 2023 16:26
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 07 00:22:34 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112723.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 28 04:28:00 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.449 min
Delta R.T.: -0.004 min
Response: 72855246
Conc: 55.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.720 min
Delta R.T.: -0.006 min
Response: 86588727
Conc: 48.12 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.242 min
Delta R.T.: -0.006 min
Response: 49225074
Conc: 47.70 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.843 min
Delta R.T.: -0.012 min
Response: 95669132
Conc: 46.35 ng/ml



#16 AR-1242-1


R.T.: 5.625 min
Delta R.T.: -0.004 min
Response: 19415926
Conc: 488.54 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.833 min
Delta R.T.: -0.006 min
Response: 24375099
Conc: 460.03 ng/ml




#17 AR-1242-2

R.T.: 5.648 min
Delta R.T.: -0.004 min
Response: 28532635
Conc: 479.91 ng/ml



#17 AR-1242-2


R.T.: 4.852 min
Delta R.T.: -0.006 min
Response: 34295449
Conc: 471.01 ng/ml



#18 AR-1242-3


R.T.: 5.710 min
Delta R.T.: -0.004 min
Response: 18839253
Conc: 481.79 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 5.033 min
Delta R.T.: -0.007 min
Response: 18971009
Conc: 468.13 ng/ml




#19 AR-1242-4

R.T.: 5.809 min
Delta R.T.: -0.004 min
Response: 14823415
Conc: 488.80 ng/ml



#19 AR-1242-4


R.T.: 5.118 min
Delta R.T.: -0.006 min
Response: 20851477
Conc: 461.99 ng/ml



#20 AR-1242-5

R.T.: 6.550 min
Delta R.T.: -0.004 min
Response: 14912982
Conc: 497.39 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.650 min
Delta R.T.: -0.007 min
Response: 25115209
Conc: 479.94 ng/ml