

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120623\
 Data File : PP062227.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Dec 2023 18:36
 Operator : YP\AJ
 Sample : 05635-02
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 M00-23-00027

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 07 00:24:47 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 28 04:28:00 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.458	3.736	22108046	26886319	16.794	14.942
2) SA Decachlor...	10.240	8.844	21196429	36250965	20.539	17.564

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120623\
Data File : PP062227.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Dec 2023 18:36
Operator : YP\AJ
Sample : 05635-02
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
M00-23-00027

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 07 00:24:47 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112723.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 28 04:28:00 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.458 min
Delta R.T.: 0.005 min
Response: 22108046
Conc: 16.79 ng/ml

Instrument :
ECD_P
ClientSampleId :
M00-23-00027




#1 Tetrachloro-m-xylene

R.T.: 3.736 min
Delta R.T.: 0.010 min
Response: 26886319
Conc: 14.94 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.240 min
Delta R.T.: -0.008 min
Response: 21196429
Conc: 20.54 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.844 min
Delta R.T.: -0.011 min
Response: 36250965
Conc: 17.56 ng/ml