

Data Path : C:\Users\eUser\Desktop\practice\PL101821\
 Data File : PL070386.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Oct 2021 21:00
 Operator : AR\AJ
 Sample : M4249-04
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 08 01:43:17 2021
 Quant Method : C:\Users\eUser\Desktop\practice\PL092821.M
 Quant Title : GC Extractables
 QLast Update : Wed Sep 29 09:19:28 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.624	5.561	9102771	23119432	18.191	17.480
28)	SA Decachlor...	10.389	11.484	13453707	21914939	19.442	18.109
Target Compounds							
7)	B delta-BHC	0.000	7.079	0	1557531	N.D.	0.805 #
9)	A Endosulfan I	7.492f	0.000	306123	0	0.458	N.D. #
10)	B gamma-Chl...	0.000	8.291	0	723074	N.D.	0.452 #
21)	B Endrin ke...	0.000	10.085	0	1884132	N.D.	1.409 #
23)	Chlordane-1	0.000	6.970f	0	111464	N.D.	2.428 #
25)	Chlordane-3	0.000	8.291	0	723074	N.D.	3.414 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : C:\Users\eUser\Desktop\practice\PL101821\
Data File : PL070386.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Oct 2021 21:00
Operator : AR\AJ
Sample : M4249-04
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 08 01:43:17 2021
Quant Method : C:\Users\eUser\Desktop\practice\PL092821.M
Quant Title : GC Extractables
QLast Update : Wed Sep 29 09:19:28 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

