

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120721\
 Data File : PP041742.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 07 Dec 2021 20:10
 Operator : AJ\MA
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 08 00:11:09 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP113021.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 30 23:30:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.743	3.748	1161695	1566955	51.518	51.585
2)	SA Decachlor...	10.635	9.002	1026045	742316	47.651	48.030
Target Compounds							
26)	L6 AR-1254-1	6.963	5.861	431077	657476	507.923	472.309
27)	L6 AR-1254-2	7.193	6.021	684989	578686	519.956	484.562
28)	L6 AR-1254-3	7.573	6.439	708454	851597	497.374	481.054
29)	L6 AR-1254-4	7.869	6.679	512120	499714	493.610	484.833
30)	L6 AR-1254-5	8.297	7.107	550354	685256	490.367	469.978



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120721\
Data File : PP041742.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 07 Dec 2021 20:10
Operator : AJ\MA
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 08 00:11:09 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP113021.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 30 23:30:32 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.743 min
Delta R.T.: -0.004 min
Response: 1161695
Conc: 51.52 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.748 min
Delta R.T.: -0.003 min
Response: 1566955
Conc: 51.59 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.635 min
Delta R.T.: -0.009 min
Response: 1026045
Conc: 47.65 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.002 min
Delta R.T.: -0.011 min
Response: 742316
Conc: 48.03 ng/ml



#26 AR-1254-1


R.T.: 6.963 min
Delta R.T.: -0.007 min
Response: 431077
Conc: 507.92 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.861 min
Delta R.T.: -0.008 min
Response: 657476
Conc: 472.31 ng/ml




#27 AR-1254-2

R.T.: 7.193 min
Delta R.T.: -0.007 min
Response: 684989
Conc: 519.96 ng/ml



#27 AR-1254-2


R.T.: 6.021 min
Delta R.T.: -0.008 min
Response: 578686
Conc: 484.56 ng/ml



#28 AR-1254-3


R.T.: 7.573 min
Delta R.T.: -0.007 min
Response: 708454
Conc: 497.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.439 min
Delta R.T.: -0.009 min
Response: 851597
Conc: 481.05 ng/ml




#29 AR-1254-4

R.T.: 7.869 min
Delta R.T.: -0.008 min
Response: 512120
Conc: 493.61 ng/ml



#29 AR-1254-4


R.T.: 6.679 min
Delta R.T.: -0.009 min
Response: 499714
Conc: 484.83 ng/ml



#30 AR-1254-5

R.T.: 8.297 min
Delta R.T.: -0.007 min
Response: 550354
Conc: 490.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 7.107 min
Delta R.T.: -0.009 min
Response: 685256
Conc: 469.98 ng/ml