

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120723\
 Data File : PP062256.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Dec 2023 13:15
 Operator : YP\AJ
 Sample : 05518-02
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 1766

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/08/2023
 Supervised By :Ankita Jodhani 12/08/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 07 14:19:34 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 28 04:28:00 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.450	3.721	30523994	37555958	23.186	20.871
2) SA Decachlor...	10.262	8.861	17534016	35844692	16.990m	17.367m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120723\
Data File : PP062256.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Dec 2023 13:15
Operator : YP\AJ
Sample : 05518-02
Misc :
ALS Vial : 11 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
1766


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/08/2023
Supervised By :Ankita Jodhani 12/08/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 07 14:19:34 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112723.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 28 04:28:00 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.450 min
Delta R.T.: -0.003 min
Response: 30523994
Conc: 23.19 ng/ml

Instrument :
ECD_P
ClientSampleId :
1766


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/08/2023
Supervised By :Ankita Jodhani 12/08/2023




#1 Tetrachloro-m-xylene

R.T.: 3.721 min
Delta R.T.: -0.005 min
Response: 37555958
Conc: 20.87 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.262 min
Delta R.T.: 0.013 min
Response: 17534016
Conc: 16.99 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.861 min
Delta R.T.: 0.005 min
Response: 35844692
Conc: 17.37 ng/ml m