

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120723\
 Data File : PP062283.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Dec 2023 23:01
 Operator : YP\AJ
 Sample : 05658-02 10X
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 34147

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/08/2023
 Supervised By :Ankita Jodhani 12/08/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 08 00:42:34 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 28 04:28:00 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.448	3.720	2701418	4176847	2.052m	2.321
2) SA Decachlor...	10.243	8.840	1501734	2345788	1.455	1.137m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120723\
Data File : PP062283.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Dec 2023 23:01
Operator : YP\AJ
Sample : 05658-02 10X
Misc :
ALS Vial : 28 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
34147


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/08/2023
Supervised By :Ankita Jodhani 12/08/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 08 00:42:34 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112723.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 28 04:28:00 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.448 min
Delta R.T.: -0.005 min
Response: 2701418
Conc: 2.05 ng/ml

Instrument :
ECD_P
ClientSampleId :
34147


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/08/2023
Supervised By :Ankita Jodhani 12/08/2023




#1 Tetrachloro-m-xylene

R.T.: 3.720 min
Delta R.T.: -0.006 min
Response: 4176847
Conc: 2.32 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.243 min
Delta R.T.: -0.005 min
Response: 1501734
Conc: 1.46 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.840 min
Delta R.T.: -0.015 min
Response: 2345788
Conc: 1.14 ng/ml m