

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120723\
 Data File : PP062294.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Dec 2023 02:31
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 08 03:14:30 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 28 04:28:00 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.449	3.720	76335077	87954151	57.985	48.880
2) SA Decachlor...	10.239	8.839	51098080	92694857	49.513	44.912
Target Compounds						
16) L4 AR-1242-1	5.624	4.832	21434529	25158084	539.332	474.808
17) L4 AR-1242-2	5.646	4.852	29492389	35464026	496.055	487.057
18) L4 AR-1242-3	5.709	5.031	19435009	19240065	497.025	474.772
19) L4 AR-1242-4	5.807	5.116	16558415	21213407	546.011	470.009
20) L4 AR-1242-5	6.548	5.648	16583051	26731929	553.094	510.833



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120723\
Data File : PP062294.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Dec 2023 02:31
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 08 03:14:30 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112723.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 28 04:28:00 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.449 min
Delta R.T.: -0.004 min
Response: 76335077
Conc: 57.99 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.720 min
Delta R.T.: -0.006 min
Response: 87954151
Conc: 48.88 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.239 min
Delta R.T.: -0.009 min
Response: 51098080
Conc: 49.51 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.839 min
Delta R.T.: -0.016 min
Response: 92694857
Conc: 44.91 ng/ml



#16 AR-1242-1


R.T.: 5.624 min
Delta R.T.: -0.005 min
Response: 21434529
Conc: 539.33 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.832 min
Delta R.T.: -0.008 min
Response: 25158084
Conc: 474.81 ng/ml




#17 AR-1242-2

R.T.: 5.646 min
Delta R.T.: -0.005 min
Response: 29492389
Conc: 496.05 ng/ml



#17 AR-1242-2


R.T.: 4.852 min
Delta R.T.: -0.007 min
Response: 35464026
Conc: 487.06 ng/ml



#18 AR-1242-3


R.T.: 5.709 min
Delta R.T.: -0.005 min
Response: 19435009
Conc: 497.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 5.031 min
Delta R.T.: -0.008 min
Response: 19240065
Conc: 474.77 ng/ml




#19 AR-1242-4

R.T.: 5.807 min
Delta R.T.: -0.005 min
Response: 16558415
Conc: 546.01 ng/ml



#19 AR-1242-4


R.T.: 5.116 min
Delta R.T.: -0.008 min
Response: 21213407
Conc: 470.01 ng/ml



#20 AR-1242-5

R.T.: 6.548 min
Delta R.T.: -0.005 min
Response: 16583051
Conc: 553.09 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.648 min
Delta R.T.: -0.009 min
Response: 26731929
Conc: 510.83 ng/ml