

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120821\
 Data File : PP041759.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 08 Dec 2021 15:47
 Operator : AJ\MA
 Sample : M4959-10
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 08 23:57:41 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP113021.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 30 23:30:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.744	3.748	497746	696027	22.073	22.914
2)	SA Decachlor...	10.638	9.004	458374	324184	21.287	20.976
Target Compounds							
8)	L2 AR-1221-1	4.996	0.000	22609	0	86.019	N.D. #
9)	L2 AR-1221-2	5.094	4.097	6188	15252	33.020	55.403 #
14)	L3 AR-1232-4	0.000	5.313	0	11714	N.D.	41.032 #
19)	L4 AR-1242-4	0.000	5.313	0	11714	N.D.	19.961 #
23)	L5 AR-1248-3	0.000	5.313	0	11714	N.D.	13.832 #
30)	L6 AR-1254-5	8.281f	7.097f	9464	11003	8.432	7.546
36)	L8 AR-1262-1	0.000	7.097f	0	11003	N.D.	14.721 #
38)	L8 AR-1262-3	0.000	7.949	0	15233	N.D.	17.093 #
41)	L9 AR-1268-1	0.000	7.949	0	15233	N.D.	6.351 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120821\
Data File : PP041759.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 08 Dec 2021 15:47
Operator : AJ\MA
Sample : M4959-10
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 08 23:57:41 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP113021.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 30 23:30:32 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.744 min
Delta R.T.: -0.003 min
Response: 497746
Conc: 22.07 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.748 min
Delta R.T.: -0.004 min
Response: 696027
Conc: 22.91 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.638 min
Delta R.T.: -0.006 min
Response: 458374
Conc: 21.29 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.004 min
Delta R.T.: -0.008 min
Response: 324184
Conc: 20.98 ng/ml



#8 AR-1221-1


R.T.: 4.996 min
Delta R.T.: -0.002 min
Response: 22609
Conc: 86.02 ng/ml

Instrument :
ECD_P
ClientSampleId :




#8 AR-1221-1

R.T.: 0.000 min
Exp R.T. : 4.008 min
Response: 0
Conc: N.D.




#9 AR-1221-2

R.T.: 5.094 min
Delta R.T.: -0.002 min
Response: 6188
Conc: 33.02 ng/ml



#9 AR-1221-2


R.T.: 4.097 min
Delta R.T.: -0.011 min
Response: 15252
Conc: 55.40 ng/ml



#14 AR-1232-4


R.T.: 0.000 min
Exp R.T.: 6.259 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#14 AR-1232-4

R.T.: 5.313 min
Delta R.T.: 0.007 min
Response: 11714
Conc: 41.03 ng/ml




#19 AR-1242-4

R.T.: 0.000 min
Exp R.T.: 6.258 min
Response: 0
Conc: N.D.



#19 AR-1242-4


R.T.: 5.313 min
Delta R.T.: 0.007 min
Response: 11714
Conc: 19.96 ng/ml



#23 AR-1248-3


R.T.: 0.000 min
Exp R.T.: 6.573 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.313 min
Delta R.T.: 0.006 min
Response: 11714
Conc: 13.83 ng/ml




#30 AR-1254-5

R.T.: 8.281 min
Delta R.T.: -0.023 min
Response: 9464
Conc: 8.43 ng/ml



#30 AR-1254-5


R.T.: 7.097 min
Delta R.T.: -0.020 min
Response: 11003
Conc: 7.55 ng/ml



#36 AR-1262-1


R.T.: 0.000 min
Exp R.T.: 8.374 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#36 AR-1262-1

R.T.: 7.097 min
Delta R.T.: -0.016 min
Response: 11003
Conc: 14.72 ng/ml




#38 AR-1262-3

R.T.: 0.000 min
Exp R.T.: 9.214 min
Response: 0
Conc: N.D.



#38 AR-1262-3


R.T.: 7.949 min
Delta R.T.: -0.002 min
Response: 15233
Conc: 17.09 ng/ml



#41 AR-1268-1

R.T.: 0.000 min
Exp R.T.: 9.214 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :



#41 AR-1268-1

R.T.: 7.949 min
Delta R.T.: -0.002 min
Response: 15233
Conc: 6.35 ng/ml