

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120823\
 Data File : PP062315.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Dec 2023 16:43
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 08 17:34:34 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 28 04:28:00 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.449	3.719	78388138	90362215	59.545	50.218
2) SA Decachlor...	10.236	8.839	49182925	87315020	47.657	42.305
Target Compounds						
16) L4 AR-1242-1	5.624	4.831	20694575	25663153	520.714	484.340
17) L4 AR-1242-2	5.646	4.851	30542148	36253556	513.712	497.900
18) L4 AR-1242-3	5.710	5.031	27676020	19558172	707.778	482.621 #
19) L4 AR-1242-4	5.810	5.116	18588187	21427279	612.943	474.748
20) L4 AR-1242-5	6.548	5.648	14575018	25724794	486.120	491.587



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120823\
Data File : PP062315.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Dec 2023 16:43
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 08 17:34:34 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112723.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 28 04:28:00 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.449 min
Delta R.T.: -0.004 min
Response: 78388138
Conc: 59.54 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.719 min
Delta R.T.: -0.007 min
Response: 90362215
Conc: 50.22 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.236 min
Delta R.T.: -0.012 min
Response: 49182925
Conc: 47.66 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.839 min
Delta R.T.: -0.017 min
Response: 87315020
Conc: 42.31 ng/ml



#16 AR-1242-1


R.T.: 5.624 min
Delta R.T.: -0.006 min
Response: 20694575
Conc: 520.71 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.831 min
Delta R.T.: -0.008 min
Response: 25663153
Conc: 484.34 ng/ml




#17 AR-1242-2

R.T.: 5.646 min
Delta R.T.: -0.006 min
Response: 30542148
Conc: 513.71 ng/ml



#17 AR-1242-2


R.T.: 4.851 min
Delta R.T.: -0.008 min
Response: 36253556
Conc: 497.90 ng/ml



#18 AR-1242-3


R.T.: 5.710 min
Delta R.T.: -0.004 min
Response: 27676020
Conc: 707.78 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 5.031 min
Delta R.T.: -0.008 min
Response: 19558172
Conc: 482.62 ng/ml




#19 AR-1242-4

R.T.: 5.810 min
Delta R.T.: -0.003 min
Response: 18588187
Conc: 612.94 ng/ml



#19 AR-1242-4


R.T.: 5.116 min
Delta R.T.: -0.008 min
Response: 21427279
Conc: 474.75 ng/ml



#20 AR-1242-5

R.T.: 6.548 min
Delta R.T.: -0.006 min
Response: 14575018
Conc: 486.12 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.648 min
Delta R.T.: -0.009 min
Response: 25724794
Conc: 491.59 ng/ml