

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120920\
 Data File : PP031958.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 10 Dec 2020 2:58
 Operator : DD\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 10 03:26:57 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112220.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 24 12:23:58 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.786	4.058	2681347	2462228	55.373	51.384
2) SA Decachlor...	10.655	9.356	1767061	1914669	52.027	51.710
Target Compounds						
26) L6 AR-1254-1	6.994	6.181	738531	1083053	484.645	500.670
27) L6 AR-1254-2	7.223	6.340	1167359	928437	502.644	498.360
28) L6 AR-1254-3	7.602	6.761	1260298	1533222	498.864	495.987
29) L6 AR-1254-4	7.897	7.000	901023	871500	531.012	512.978
30) L6 AR-1254-5	8.324	7.427	982355	1325896	525.489	491.149



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120920\
Data File : PP031958.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 10 Dec 2020 2:58
Operator : DD\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleID :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 10 03:26:57 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112220.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 24 12:23:58 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.786 min
Delta R.T.: -0.004 min
Response: 2681347
Conc: 55.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.058 min
Delta R.T.: -0.004 min
Response: 2462228
Conc: 51.38 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.655 min
Delta R.T.: -0.010 min
Response: 1767061
Conc: 52.03 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.356 min
Delta R.T.: -0.015 min
Response: 1914669
Conc: 51.71 ng/ml



#26 AR-1254-1


R.T.: 6.994 min
Delta R.T.: -0.003 min
Response: 738531
Conc: 484.64 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 6.181 min
Delta R.T.: -0.009 min
Response: 1083053
Conc: 500.67 ng/ml




#27 AR-1254-2

R.T.: 7.223 min
Delta R.T.: -0.005 min
Response: 1167359
Conc: 502.64 ng/ml



#27 AR-1254-2


R.T.: 6.340 min
Delta R.T.: -0.008 min
Response: 928437
Conc: 498.36 ng/ml



#28 AR-1254-3


R.T.: 7.602 min
Delta R.T.: -0.004 min
Response: 1260298
Conc: 498.86 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.761 min
Delta R.T.: -0.008 min
Response: 1533222
Conc: 495.99 ng/ml




#29 AR-1254-4

R.T.: 7.897 min
Delta R.T.: -0.004 min
Response: 901023
Conc: 531.01 ng/ml



#29 AR-1254-4


R.T.: 7.000 min
Delta R.T.: -0.008 min
Response: 871500
Conc: 512.98 ng/ml



#30 AR-1254-5

R.T.: 8.324 min
Delta R.T.: -0.004 min
Response: 982355
Conc: 525.49 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 7.427 min
Delta R.T.: -0.009 min
Response: 1325896
Conc: 491.15 ng/ml