

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120921\
 Data File : PP041809.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 09 Dec 2021 9:38
 Operator : AJ\MA
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 10 01:54:07 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP113021.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 30 23:30:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.744	3.746	1232755	1543164	54.669	50.802
2)	SA Decachlor...	10.636	8.997f	1096577	723324	50.926	46.801
Target Compounds							
3)	L1 AR-1016-1	6.059	4.990	405747	452369	507.097	486.923
4)	L1 AR-1016-2	6.083	5.009	635889	694974	533.702	489.561
5)	L1 AR-1016-3	6.149	5.200	408586	383586	548.307	477.766
6)	L1 AR-1016-4	6.255	5.253	327692	312306	539.795	495.629
7)	L1 AR-1016-5	6.568	5.478	305460	339625	521.265	456.573
8)	L2 AR-1221-1	4.990	4.000	43511	55653	165.541	158.584
9)	L2 AR-1221-2	5.093	4.099	55799	84732	297.746	307.790
10)	L2 AR-1221-3	5.180	4.187	225658	322434	361.872	371.531
11)	L3 AR-1232-1	5.180	4.187	225658	322434	436.688	435.281
12)	L3 AR-1232-2	5.764	5.009	332677	694974	1292.837	1119.862
13)	L3 AR-1232-3	6.083	5.200	635889	383586	1284.986	1184.290
14)	L3 AR-1232-4	6.255	5.296	327692	360564	1334.800	1263.017
15)	L3 AR-1232-5	6.352	5.478	250791	339625	1557.844	1103.752 #
16)	L4 AR-1242-1	6.059	4.990	405747	452369	676.536	614.259
17)	L4 AR-1242-2	6.083	5.009	635889	694974	692.597	623.956
18)	L4 AR-1242-3	6.149	5.200	408586	383586	703.773	624.543
19)	L4 AR-1242-4	6.255	5.296	327692	360564	699.290	614.414
20)	L4 AR-1242-5	7.035	5.857	87675	261712	182.444	402.121 #
21)	L5 AR-1248-1	6.059	5.009	405747	694974	926.840	1271.664 #
22)	L5 AR-1248-2	6.352	5.253	250791	312306	404.400	383.602
23)	L5 AR-1248-3	6.568	5.296	305460	360564	408.860	425.761
24)	L5 AR-1248-4	6.996	5.478	78928	339625	94.226	348.255 #
25)	L5 AR-1248-5	7.035	5.900	87675	43313	105.359	48.374 #
26)	L6 AR-1254-1	6.964	5.857	213327	261712	251.355	188.005 #
27)	L6 AR-1254-2	7.193	6.017	271044	231804	205.742	194.101
28)	L6 AR-1254-3	7.575	6.453	173043	461197	121.486	260.523 #
29)	L6 AR-1254-4	7.869	6.676	138321	66417	133.321	64.439 #
30)	L6 AR-1254-5	8.297	7.103	736598	669464	656.310	459.148 #
31)	L7 AR-1260-1	7.740	6.572	572708	593793	580.138	518.398
32)	L7 AR-1260-2	8.006	6.771	642687	667304	511.526	508.361
33)	L7 AR-1260-3	8.370	6.924	493795	579285	545.313	433.086
34)	L7 AR-1260-4	8.600	7.406	567103	473249	529.941	480.629
35)	L7 AR-1260-5	8.914	7.656	1058843	1021169	511.031	494.036
36)	L8 AR-1262-1	8.370	7.103	493795	669464	388.868	895.706 #
37)	L8 AR-1262-2	8.914	7.656	1058843	1021169	468.651	484.008
38)	L8 AR-1262-3	9.210	7.939	230512	239847	212.380	269.138 #
39)	L8 AR-1262-4	9.281f	8.003	477229	741219	697.262	461.532 #
40)	L8 AR-1262-5	9.931	8.504	336258	202233	375.826	278.565 #
41)	L9 AR-1268-1	9.210	7.939	230512	239847	85.168	100.007

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120921\
 Data File : PP041809.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 09 Dec 2021 9:38
 Operator : AJ\MA
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 10 01:54:07 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP113021.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 30 23:30:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	9.281f	8.003	477229	741219	182.239	327.300 #
43) L9	AR-1268-3	9.511	8.208	12213	10401	5.511	5.538
44) L9	AR-1268-4	9.931	8.504	336258	202233	342.928	252.343 #
45) L9	AR-1268-5	10.322	8.774	107028	70011	14.806	13.687



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP120921\
Data File : PP041809.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 09 Dec 2021 9:38
Operator : AJ\MA
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 10 01:54:07 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP113021.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 30 23:30:32 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.744 min
Delta R.T.: -0.003 min
Response: 1232755
Conc: 54.67 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.746 min
Delta R.T.: -0.006 min
Response: 1543164
Conc: 50.80 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.636 min
Delta R.T.: -0.008 min
Response: 1096577
Conc: 50.93 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.997 min
Delta R.T.: -0.015 min
Response: 723324
Conc: 46.80 ng/ml



#3 AR-1016-1


R.T.: 6.059 min
Delta R.T.: -0.004 min
Response: 405747
Conc: 507.10 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.990 min
Delta R.T.: -0.008 min
Response: 452369
Conc: 486.92 ng/ml




#4 AR-1016-2

R.T.: 6.083 min
Delta R.T.: -0.004 min
Response: 635889
Conc: 533.70 ng/ml



#4 AR-1016-2


R.T.: 5.009 min
Delta R.T.: -0.009 min
Response: 694974
Conc: 489.56 ng/ml



#5 AR-1016-3


R.T.: 6.149 min
Delta R.T.: -0.003 min
Response: 408586
Conc: 548.31 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 5.200 min
Delta R.T.: -0.009 min
Response: 383586
Conc: 477.77 ng/ml




#6 AR-1016-4

R.T.: 6.255 min
Delta R.T.: -0.004 min
Response: 327692
Conc: 539.80 ng/ml



#6 AR-1016-4


R.T.: 5.253 min
Delta R.T.: -0.009 min
Response: 312306
Conc: 495.63 ng/ml



#7 AR-1016-5


R.T.: 6.568 min
Delta R.T.: -0.004 min
Response: 305460
Conc: 521.27 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.478 min
Delta R.T.: -0.011 min
Response: 339625
Conc: 456.57 ng/ml




#8 AR-1221-1

R.T.: 4.990 min
Delta R.T.: -0.008 min
Response: 43511
Conc: 165.54 ng/ml



#8 AR-1221-1


R.T.: 4.000 min
Delta R.T.: -0.009 min
Response: 55653
Conc: 158.58 ng/ml



#9 AR-1221-2


R.T.: 5.093 min
Delta R.T.: -0.004 min
Response: 55799
Conc: 297.75 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 4.099 min
Delta R.T.: -0.008 min
Response: 84732
Conc: 307.79 ng/ml




#10 AR-1221-3

R.T.: 5.180 min
Delta R.T.: -0.003 min
Response: 225658
Conc: 361.87 ng/ml



#10 AR-1221-3


R.T.: 4.187 min
Delta R.T.: -0.008 min
Response: 322434
Conc: 371.53 ng/ml



#11 AR-1232-1


R.T.: 5.180 min
Delta R.T.: -0.002 min
Response: 225658
Conc: 436.69 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 4.187 min
Delta R.T.: -0.008 min
Response: 322434
Conc: 435.28 ng/ml




#12 AR-1232-2

R.T.: 5.764 min
Delta R.T.: -0.004 min
Response: 332677
Conc: 1292.84 ng/ml



#12 AR-1232-2


R.T.: 5.009 min
Delta R.T.: -0.009 min
Response: 694974
Conc: 1119.86 ng/ml



#13 AR-1232-3


R.T.: 6.083 min
Delta R.T.: -0.004 min
Response: 635889
Conc: 1284.99 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 5.200 min
Delta R.T.: -0.009 min
Response: 383586
Conc: 1184.29 ng/ml




#14 AR-1232-4

R.T.: 6.255 min
Delta R.T.: -0.004 min
Response: 327692
Conc: 1334.80 ng/ml



#14 AR-1232-4


R.T.: 5.296 min
Delta R.T.: -0.009 min
Response: 360564
Conc: 1263.02 ng/ml



#15 AR-1232-5


R.T.: 6.352 min
Delta R.T.: -0.005 min
Response: 250791
Conc: 1557.84 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.478 min
Delta R.T.: -0.011 min
Response: 339625
Conc: 1103.75 ng/ml




#16 AR-1242-1

R.T.: 6.059 min
Delta R.T.: -0.003 min
Response: 405747
Conc: 676.54 ng/ml



#16 AR-1242-1


R.T.: 4.990 min
Delta R.T.: -0.009 min
Response: 452369
Conc: 614.26 ng/ml



#17 AR-1242-2


R.T.: 6.083 min
Delta R.T.: -0.003 min
Response: 635889
Conc: 692.60 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 5.009 min
Delta R.T.: -0.009 min
Response: 694974
Conc: 623.96 ng/ml




#18 AR-1242-3

R.T.: 6.149 min
Delta R.T.: -0.004 min
Response: 408586
Conc: 703.77 ng/ml



#18 AR-1242-3


R.T.: 5.200 min
Delta R.T.: -0.009 min
Response: 383586
Conc: 624.54 ng/ml



#19 AR-1242-4


R.T.: 6.255 min
Delta R.T.: -0.003 min
Response: 327692
Conc: 699.29 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 5.296 min
Delta R.T.: -0.010 min
Response: 360564
Conc: 614.41 ng/ml




#20 AR-1242-5

R.T.: 7.035 min
Delta R.T.: -0.004 min
Response: 87675
Conc: 182.44 ng/ml



#20 AR-1242-5


R.T.: 5.857 min
Delta R.T.: -0.011 min
Response: 261712
Conc: 402.12 ng/ml



#21 AR-1248-1


R.T.: 6.059 min
Delta R.T.: -0.005 min
Response: 405747
Conc: 926.84 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 5.009 min
Delta R.T.: 0.008 min
Response: 694974
Conc: 1271.66 ng/ml




#22 AR-1248-2

R.T.: 6.352 min
Delta R.T.: -0.006 min
Response: 250791
Conc: 404.40 ng/ml



#22 AR-1248-2


R.T.: 5.253 min
Delta R.T.: -0.011 min
Response: 312306
Conc: 383.60 ng/ml



#23 AR-1248-3


R.T.: 6.568 min
Delta R.T.: -0.005 min
Response: 305460
Conc: 408.86 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.296 min
Delta R.T.: -0.011 min
Response: 360564
Conc: 425.76 ng/ml




#24 AR-1248-4

R.T.: 6.996 min
Delta R.T.: -0.006 min
Response: 78928
Conc: 94.23 ng/ml



#24 AR-1248-4


R.T.: 5.478 min
Delta R.T.: -0.012 min
Response: 339625
Conc: 348.26 ng/ml



#25 AR-1248-5


R.T.: 7.035 min
Delta R.T.: -0.006 min
Response: 87675
Conc: 105.36 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.900 min
Delta R.T.: -0.012 min
Response: 43313
Conc: 48.37 ng/ml




#26 AR-1254-1

R.T.: 6.964 min
Delta R.T.: -0.006 min
Response: 213327
Conc: 251.36 ng/ml



#26 AR-1254-1


R.T.: 5.857 min
Delta R.T.: -0.012 min
Response: 261712
Conc: 188.00 ng/ml



#27 AR-1254-2


R.T.: 7.193 min
Delta R.T.: -0.007 min
Response: 271044
Conc: 205.74 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 6.017 min
Delta R.T.: -0.012 min
Response: 231804
Conc: 194.10 ng/ml




#28 AR-1254-3

R.T.: 7.575 min
Delta R.T.: -0.006 min
Response: 173043
Conc: 121.49 ng/ml



#28 AR-1254-3


R.T.: 6.453 min
Delta R.T.: 0.005 min
Response: 461197
Conc: 260.52 ng/ml



#29 AR-1254-4


R.T.: 7.869 min
Delta R.T.: -0.008 min
Response: 138321
Conc: 133.32 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.676 min
Delta R.T.: -0.012 min
Response: 66417
Conc: 64.44 ng/ml




#30 AR-1254-5

R.T.: 8.297 min
Delta R.T.: -0.007 min
Response: 736598
Conc: 656.31 ng/ml



#30 AR-1254-5


R.T.: 7.103 min
Delta R.T.: -0.014 min
Response: 669464
Conc: 459.15 ng/ml



#31 AR-1260-1


R.T.: 7.740 min
Delta R.T.: -0.005 min
Response: 572708
Conc: 580.14 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.572 min
Delta R.T.: -0.011 min
Response: 593793
Conc: 518.40 ng/ml




#32 AR-1260-2

R.T.: 8.006 min
Delta R.T.: -0.006 min
Response: 642687
Conc: 511.53 ng/ml



#32 AR-1260-2


R.T.: 6.771 min
Delta R.T.: -0.011 min
Response: 667304
Conc: 508.36 ng/ml



#33 AR-1260-3


R.T.: 8.370 min
Delta R.T.: -0.005 min
Response: 493795
Conc: 545.31 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.924 min
Delta R.T.: -0.011 min
Response: 579285
Conc: 433.09 ng/ml




#34 AR-1260-4

R.T.: 8.600 min
Delta R.T.: -0.006 min
Response: 567103
Conc: 529.94 ng/ml



#34 AR-1260-4


R.T.: 7.406 min
Delta R.T.: -0.012 min
Response: 473249
Conc: 480.63 ng/ml



#35 AR-1260-5


R.T.: 8.914 min
Delta R.T.: -0.007 min
Response: 1058843
Conc: 511.03 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.656 min
Delta R.T.: -0.012 min
Response: 1021169
Conc: 494.04 ng/ml




#36 AR-1262-1

R.T.: 8.370 min
Delta R.T.: -0.004 min
Response: 493795
Conc: 388.87 ng/ml



#36 AR-1262-1


R.T.: 7.103 min
Delta R.T.: -0.010 min
Response: 669464
Conc: 895.71 ng/ml



#37 AR-1262-2


R.T.: 8.914 min
Delta R.T.: -0.006 min
Response: 1058843
Conc: 468.65 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 7.656 min
Delta R.T.: -0.010 min
Response: 1021169
Conc: 484.01 ng/ml




#38 AR-1262-3

R.T.: 9.210 min
Delta R.T.: -0.004 min
Response: 230512
Conc: 212.38 ng/ml



#38 AR-1262-3


R.T.: 7.939 min
Delta R.T.: -0.012 min
Response: 239847
Conc: 269.14 ng/ml



#39 AR-1262-4


R.T.: 9.281 min
Delta R.T.: -0.022 min
Response: 477229
Conc: 697.26 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 8.003 min
Delta R.T.: -0.012 min
Response: 741219
Conc: 461.53 ng/ml




#40 AR-1262-5

R.T.: 9.931 min
Delta R.T.: -0.006 min
Response: 336258
Conc: 375.83 ng/ml



#40 AR-1262-5


R.T.: 8.504 min
Delta R.T.: -0.010 min
Response: 202233
Conc: 278.57 ng/ml



#41 AR-1268-1


R.T.: 9.210 min
Delta R.T.: -0.004 min
Response: 230512
Conc: 85.17 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.939 min
Delta R.T.: -0.011 min
Response: 239847
Conc: 100.01 ng/ml




#42 AR-1268-2

R.T.: 9.281 min
Delta R.T.: -0.025 min
Response: 477229
Conc: 182.24 ng/ml



#42 AR-1268-2


R.T.: 8.003 min
Delta R.T.: -0.013 min
Response: 741219
Conc: 327.30 ng/ml



#43 AR-1268-3


R.T.: 9.511 min
Delta R.T.: -0.007 min
Response: 12213
Conc: 5.51 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 8.208 min
Delta R.T.: -0.013 min
Response: 10401
Conc: 5.54 ng/ml




#44 AR-1268-4

R.T.: 9.931 min
Delta R.T.: -0.007 min
Response: 336258
Conc: 342.93 ng/ml



#44 AR-1268-4


R.T.: 8.504 min
Delta R.T.: -0.011 min
Response: 202233
Conc: 252.34 ng/ml



#45 AR-1268-5

R.T.: 10.322 min
Delta R.T.: -0.005 min
Response: 107028
Conc: 14.81 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.774 min
Delta R.T.: -0.012 min
Response: 70011
Conc: 13.69 ng/ml