

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121024\
 Data File : PP068907.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Dec 2024 18:34
 Operator : YP\AJ
 Sample : AR1268ICC050
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/11/2024
 Supervised By :Ankita Jodhani 12/11/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 10 23:53:11 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121024.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Dec 10 23:51:55 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.668	3.981	7185437	4811563	4.901	4.627
2) SA Decachlor...	10.510	9.105	9868722	10696380	5.086	5.475
Target Compounds						
41) L9 AR-1268-1	8.931	7.921	9781508	8519337	53.336m	53.433m
42) L9 AR-1268-2	9.030	7.986	8152120	7630193	48.740	52.172
43) L9 AR-1268-3	9.273	8.200	7465470	6794612	52.438	52.285
44) L9 AR-1268-4	9.710	8.504	2755871	2673620	47.850	47.534
45) L9 AR-1268-5	10.148	8.823	19666427	19276746	50.495m	50.668

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121024\
Data File : PP068907.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Dec 2024 18:34
Operator : YP\AJ
Sample : AR1268ICC050
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/11/2024
Supervised By :Ankita Jodhani 12/11/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 10 23:53:11 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121024.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Dec 10 23:51:55 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

