

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121025\
 Data File : PP076928.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Dec 2025 13:01
 Operator : YP\AJ
 Sample : Q3823-01 10X
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 BUR-1321

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 12/11/2025
 Supervised By :Yogesh Patel 12/11/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 11 00:43:24 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Nov 21 22:52:01 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.619	3.771	2652297	2464553	1.445m	1.183
2) SA Decachlor...	10.380	8.764	4703945	3018628	3.139	1.711 #
Target Compounds						
26) L6 AR-1254-1	6.624	5.649	2755033	3342286	34.344m	28.267
27) L6 AR-1254-2	6.838	5.795	4170905	4666539	37.000m	43.159
28) L6 AR-1254-3	7.200	6.196	7404433	7719254	61.629m	47.015
29) L6 AR-1254-4	7.482	6.423	6303618	5092702	69.948m	49.066 #
30) L6 AR-1254-5	7.898	6.839	8261124	11087297	85.347m	78.759

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121025\
Data File : PP076928.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Dec 2025 13:01
Operator : YP\AJ
Sample : Q3823-01 10X
Misc :
ALS Vial : 8 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
BUR-1321


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 12/11/2025
Supervised By :Yogesh Patel 12/11/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 11 00:43:24 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Nov 21 22:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.619 min
Delta R.T.: 0.000 min
Response: 2652297
Conc: 1.44 ng/ml

Instrument :
ECD_P
ClientSampleId :
BUR-1321


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 12/11/2025
Supervised By :Yogesh Patel 12/11/2025




#1 Tetrachloro-m-xylene

R.T.: 3.771 min
Delta R.T.: -0.004 min
Response: 2464553
Conc: 1.18 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.380 min
Delta R.T.: 0.011 min
Response: 4703945
Conc: 3.14 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.764 min
Delta R.T.: -0.007 min
Response: 3018628
Conc: 1.71 ng/ml




#26 AR-1254-1

R.T.: 6.624 min
Delta R.T.: 0.009 min
Response: 2755033
Conc: 34.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
BUR-1321


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 12/11/2025
Supervised By :Yogesh Patel 12/11/2025




#26 AR-1254-1

R.T.: 5.649 min
Delta R.T.: 0.001 min
Response: 3342286
Conc: 28.27 ng/ml




#27 AR-1254-2

R.T.: 6.838 min
Delta R.T.: 0.006 min
Response: 4170905
Conc: 37.00 ng/ml m



#27 AR-1254-2

R.T.: 5.795 min
Delta R.T.: 0.000 min
Response: 4666539
Conc: 43.16 ng/ml




#28 AR-1254-3

R.T.: 7.200 min
Delta R.T.: 0.007 min
Response: 7404433
Conc: 61.63 ng/ml

Instrument :
ECD_P
ClientSampleId :
BUR-1321


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 12/11/2025
Supervised By :Yogesh Patel 12/11/2025




#28 AR-1254-3

R.T.: 6.196 min
Delta R.T.: 0.000 min
Response: 7719254
Conc: 47.01 ng/ml




#29 AR-1254-4

R.T.: 7.482 min
Delta R.T.: 0.008 min
Response: 6303618
Conc: 69.95 ng/ml m



#29 AR-1254-4

R.T.: 6.423 min
Delta R.T.: 0.000 min
Response: 5092702
Conc: 49.07 ng/ml




#30 AR-1254-5

R.T.: 7.898 min
Delta R.T.: 0.008 min
Response: 8261124
Conc: 85.35 ng/ml

Instrument :
ECD_P
ClientSampleId :
BUR-1321

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 12/11/2025
Supervised By :Yogesh Patel 12/11/2025



#30 AR-1254-5

R.T.: 6.839 min
Delta R.T.: 0.000 min
Response: 11087297
Conc: 78.76 ng/ml