

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121123\
 Data File : PP062389.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Dec 2023 02:08
 Operator : YP\AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 12 04:08:24 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Dec 12 04:05:44 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.434	3.719	87554974	155.7E6	51.563	51.073
2) SA Decachlor...	10.209	8.836	61658393	129.8E6	52.013	53.829
Target Compounds						
26) L6 AR-1254-1	6.466	5.647	34839802	80481589	515.463	509.172
27) L6 AR-1254-2	6.686	5.797	53645167	71421791	518.897	508.221
28) L6 AR-1254-3	7.053	6.208	56157191	112.3E6	521.224	507.156
29) L6 AR-1254-4	7.340	6.438	35126106	56815394	543.717	510.565
30) L6 AR-1254-5	7.760	6.862	37858642	95024326	509.051	508.739



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121123\
Data File : PP062389.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Dec 2023 02:08
Operator : YP\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 12 04:08:24 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121123.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Dec 12 04:05:44 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.434 min
Delta R.T.: 0.000 min
Response: 87554974
Conc: 51.56 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.719 min
Delta R.T.: 0.000 min
Response: 155689000
Conc: 51.07 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.209 min
Delta R.T.: 0.000 min
Response: 61658393
Conc: 52.01 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.836 min
Delta R.T.: 0.000 min
Response: 129820782
Conc: 53.83 ng/ml



#26 AR-1254-1


R.T.: 6.466 min
Delta R.T.: 0.000 min
Response: 34839802
Conc: 515.46 ng/ml

Instrument :
ECD_P
ClientSampleId :




#26 AR-1254-1

R.T.: 5.647 min
Delta R.T.: 0.000 min
Response: 80481589
Conc: 509.17 ng/ml




#27 AR-1254-2

R.T.: 6.686 min
Delta R.T.: 0.000 min
Response: 53645167
Conc: 518.90 ng/ml



#27 AR-1254-2


R.T.: 5.797 min
Delta R.T.: 0.000 min
Response: 71421791
Conc: 508.22 ng/ml



#28 AR-1254-3


R.T.: 7.053 min
Delta R.T.: 0.000 min
Response: 56157191
Conc: 521.22 ng/ml

Instrument :
ECD_P
ClientSampleId :




#28 AR-1254-3

R.T.: 6.208 min
Delta R.T.: 0.000 min
Response: 112270932
Conc: 507.16 ng/ml




#29 AR-1254-4

R.T.: 7.340 min
Delta R.T.: 0.000 min
Response: 35126106
Conc: 543.72 ng/ml



#29 AR-1254-4


R.T.: 6.438 min
Delta R.T.: 0.000 min
Response: 56815394
Conc: 510.56 ng/ml



#30 AR-1254-5

R.T.: 7.760 min
Delta R.T.: -0.001 min
Response: 37858642
Conc: 509.05 ng/ml

Instrument :
ECD_P
ClientSampleId :



#30 AR-1254-5

R.T.: 6.862 min
Delta R.T.: 0.000 min
Response: 95024326
Conc: 508.74 ng/ml