

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121222\
 Data File : PP053692.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Dec 2022 20:30
 Operator : YP\AJ
 Sample : N5996-04 10X
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 BC271299-2-2

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 12 23:32:40 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP120822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Dec 08 18:35:17 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.339	3.594	3515014	2531992	2.066	1.696
2)	SA Decachlor...	10.101	8.639	3962829	2215912	3.265	2.007 #
Target Compounds							
26)	L6 AR-1254-1	6.370	5.494	8207753	7600956	124.054	117.126
27)	L6 AR-1254-2	6.590	5.642	12095602	7165594	122.086	124.761
28)	L6 AR-1254-3	6.957	6.048	11788067	9298459	119.429	112.198
29)	L6 AR-1254-4	7.244	6.278	8404154	4968559	135.654	131.172
30)	L6 AR-1254-5	7.665	6.699	9351214	7507120	124.527	112.144

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121222\
Data File : PP053692.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Dec 2022 20:30
Operator : YP\AJ
Sample : N5996-04 10X
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
BC271299-2-2

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 12 23:32:40 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP120822.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Dec 08 18:35:17 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

