

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121222\
 Data File : PP053691.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Dec 2022 20:13
 Operator : YP\AJ
 Sample : N5996-03 10X
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 BC271299-2-1

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/13/2022
 Supervised By :Ankita Jodhani 12/13/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 12 23:31:45 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP120822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Dec 08 18:35:17 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.340	3.595	3473555	2284685	2.042	1.530 #
2) SA Decachlor...	10.101	8.640	3655099	2109530	3.012m	1.910 #
Target Compounds						
26) L6 AR-1254-1	6.371	5.494	8338986	7220713	126.038	111.267
27) L6 AR-1254-2	6.590	5.643	13176606	6977117	132.997	121.480
28) L6 AR-1254-3	6.958	6.049	12363911	9309001	125.263	112.326
29) L6 AR-1254-4	7.245	6.279	8958724	5134774	144.605	135.560
30) L6 AR-1254-5	7.665	6.700	9801014	7564550	130.517	113.002

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121222\
Data File : PP053691.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Dec 2022 20:13
Operator : YP\AJ
Sample : N5996-03 10X
Misc :
ALS Vial : 21 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
BC271299-2-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/13/2022
Supervised By :Ankita Jodhani 12/13/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 12 23:31:45 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP120822.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Dec 08 18:35:17 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.340 min
Delta R.T.: -0.002 min
Response: 3473555
Conc: 2.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
BC271299-2-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/13/2022
Supervised By :Ankita Jodhani 12/13/2022




#1 Tetrachloro-m-xylene

R.T.: 3.595 min
Delta R.T.: -0.004 min
Response: 2284685
Conc: 1.53 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.101 min
Delta R.T.: -0.001 min
Response: 3655099
Conc: 3.01 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.640 min
Delta R.T.: -0.005 min
Response: 2109530
Conc: 1.91 ng/ml




#26 AR-1254-1

R.T.: 6.371 min
Delta R.T.: 0.000 min
Response: 8338986
Conc: 126.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
BC271299-2-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/13/2022
Supervised By :Ankita Jodhani 12/13/2022




#26 AR-1254-1

R.T.: 5.494 min
Delta R.T.: -0.005 min
Response: 7220713
Conc: 111.27 ng/ml




#27 AR-1254-2

R.T.: 6.590 min
Delta R.T.: 0.000 min
Response: 13176606
Conc: 133.00 ng/ml



#27 AR-1254-2

R.T.: 5.643 min
Delta R.T.: -0.005 min
Response: 6977117
Conc: 121.48 ng/ml




#28 AR-1254-3

R.T.: 6.958 min
Delta R.T.: 0.000 min
Response: 12363911
Conc: 125.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
BC271299-2-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/13/2022
Supervised By :Ankita Jodhani 12/13/2022




#28 AR-1254-3

R.T.: 6.049 min
Delta R.T.: -0.004 min
Response: 9309001
Conc: 112.33 ng/ml




#29 AR-1254-4

R.T.: 7.245 min
Delta R.T.: 0.000 min
Response: 8958724
Conc: 144.61 ng/ml



#29 AR-1254-4

R.T.: 6.279 min
Delta R.T.: -0.004 min
Response: 5134774
Conc: 135.56 ng/ml




#30 AR-1254-5

R.T.: 7.665 min
Delta R.T.: 0.000 min
Response: 9801014
Conc: 130.52 ng/ml

Instrument :
ECD_P
ClientSampleId :
BC271299-2-1

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/13/2022
Supervised By :Ankita Jodhani 12/13/2022



#30 AR-1254-5

R.T.: 6.700 min
Delta R.T.: -0.004 min
Response: 7564550
Conc: 113.00 ng/ml