

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121222\
 Data File : PP053693.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Dec 2022 20:46
 Operator : YP\AJ
 Sample : N5996-05 10X
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 BC255396-1-1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 12 23:33:35 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP120822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Dec 08 18:35:17 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.339	3.594	3294249	2388336	1.936	1.600
2) SA Decachlor...	10.101	8.639	3629244	2284148	2.991	2.069 #

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121222\
Data File : PP053693.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Dec 2022 20:46
Operator : YP\AJ
Sample : N5996-05 10X
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
BC255396-1-1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 12 23:33:35 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP120822.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Dec 08 18:35:17 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.339 min
Delta R.T.: -0.003 min
Response: 3294249
Conc: 1.94 ng/ml

Instrument :
ECD_P
ClientSampleId :
BC255396-1-1




#1 Tetrachloro-m-xylene

R.T.: 3.594 min
Delta R.T.: -0.004 min
Response: 2388336
Conc: 1.60 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.101 min
Delta R.T.: -0.001 min
Response: 3629244
Conc: 2.99 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.639 min
Delta R.T.: -0.006 min
Response: 2284148
Conc: 2.07 ng/ml