

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121223\
 Data File : PP062393.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Dec 2023 10:52
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 12 23:23:51 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Dec 12 05:13:02 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.432	3.718	79192248	138.8E6	47.209	45.241
2) SA Decachlor...	10.204	8.833	56509667	115.6E6	54.076	50.216
Target Compounds						
16) L4 AR-1242-1	5.607	4.829	19439063	33927204	471.736	452.080
17) L4 AR-1242-2	5.630	4.848	29199657	48333580	469.407	452.706
18) L4 AR-1242-3	5.692	5.028	19534041	26637713	481.892	457.261
19) L4 AR-1242-4	5.790	5.113	15813265	29016538	468.749	455.443
20) L4 AR-1242-5	6.530	5.645	17139018	32787141	481.804	470.220



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121223\
Data File : PP062393.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Dec 2023 10:52
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 12 23:23:51 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121123.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Dec 12 05:13:02 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.432 min
Delta R.T.: -0.002 min
Response: 79192248
Conc: 47.21 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.718 min
Delta R.T.: -0.002 min
Response: 138835072
Conc: 45.24 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.204 min
Delta R.T.: -0.008 min
Response: 56509667
Conc: 54.08 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.833 min
Delta R.T.: -0.005 min
Response: 115647430
Conc: 50.22 ng/ml



#16 AR-1242-1


R.T.: 5.607 min
Delta R.T.: -0.001 min
Response: 19439063
Conc: 471.74 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.829 min
Delta R.T.: -0.002 min
Response: 33927204
Conc: 452.08 ng/ml




#17 AR-1242-2

R.T.: 5.630 min
Delta R.T.: 0.000 min
Response: 29199657
Conc: 469.41 ng/ml



#17 AR-1242-2


R.T.: 4.848 min
Delta R.T.: -0.002 min
Response: 48333580
Conc: 452.71 ng/ml



#18 AR-1242-3


R.T.: 5.692 min
Delta R.T.: -0.001 min
Response: 19534041
Conc: 481.89 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 5.028 min
Delta R.T.: -0.002 min
Response: 26637713
Conc: 457.26 ng/ml




#19 AR-1242-4

R.T.: 5.790 min
Delta R.T.: -0.002 min
Response: 15813265
Conc: 468.75 ng/ml



#19 AR-1242-4


R.T.: 5.113 min
Delta R.T.: -0.003 min
Response: 29016538
Conc: 455.44 ng/ml



#20 AR-1242-5

R.T.: 6.530 min
Delta R.T.: -0.002 min
Response: 17139018
Conc: 481.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.645 min
Delta R.T.: -0.003 min
Response: 32787141
Conc: 470.22 ng/ml