

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121223\
 Data File : PP062400.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Dec 2023 12:46
 Operator : YP\AJ
 Sample : 05763-02
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 02-A-02-B-02-C

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/13/2023
 Supervised By :Ankita Jodhani 12/13/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 12 23:25:53 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Dec 12 05:13:02 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.432	3.717	36426181	60285210	21.715	19.645
2) SA Decachlor...	10.205	8.833	24556163	46733720	23.499	20.292
Target Compounds						
26) L6 AR-1254-1	6.468	5.651	21943889	42735230	323.649m	267.965
27) L6 AR-1254-2	6.687	5.796	20658298	14561830	195.982	102.920 #
28) L6 AR-1254-3	7.051	6.206	10466792	17780618	96.678	80.219
29) L6 AR-1254-4	7.336	6.437	4014740	5457501	62.529	49.504
30) L6 AR-1254-5	7.757	6.859	4835280	21291313	65.956	114.738 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121223\
Data File : PP062400.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Dec 2023 12:46
Operator : YP\AJ
Sample : 05763-02
Misc :
ALS Vial : 10 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
02-A-02-B-02-C


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/13/2023
Supervised By :Ankita Jodhani 12/13/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 12 23:25:53 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121123.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Dec 12 05:13:02 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





#1 Tetrachloro-m-xylene

R.T.: 4.432 min
Delta R.T.: -0.002 min
Response: 36426181
Conc: 21.71 ng/ml

Instrument :

ECD_P


ClientSampleId :

02-A-02-B-02-C

Manual Integrations


APPROVED

Reviewed By :Yogesh Patel 12/13/2023
Supervised By :Ankita Jodhani 12/13/2023




#1 Tetrachloro-m-xylene

R.T.: 3.717 min
Delta R.T.: -0.003 min
Response: 60285210
Conc: 19.64 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.205 min
Delta R.T.: -0.008 min
Response: 24556163
Conc: 23.50 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.833 min
Delta R.T.: -0.005 min
Response: 46733720
Conc: 20.29 ng/ml




#26 AR-1254-1

R.T.: 6.468 min
Delta R.T.: 0.002 min
Response: 21943889
Conc: 323.65 ng/ml

Instrument :
ECD_P
ClientSampleId :
02-A-02-B-02-C


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/13/2023
Supervised By :Ankita Jodhani 12/13/2023




#26 AR-1254-1

R.T.: 5.651 min
Delta R.T.: 0.002 min
Response: 42735230
Conc: 267.97 ng/ml




#27 AR-1254-2

R.T.: 6.687 min
Delta R.T.: 0.000 min
Response: 20658298
Conc: 195.98 ng/ml



#27 AR-1254-2

R.T.: 5.796 min
Delta R.T.: -0.002 min
Response: 14561830
Conc: 102.92 ng/ml




#28 AR-1254-3

R.T.: 7.051 min
Delta R.T.: -0.003 min
Response: 10466792
Conc: 96.68 ng/ml

Instrument :
ECD_P
ClientSampleId :
02-A-02-B-02-C


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/13/2023
Supervised By :Ankita Jodhani 12/13/2023




#28 AR-1254-3

R.T.: 6.206 min
Delta R.T.: -0.002 min
Response: 17780618
Conc: 80.22 ng/ml




#29 AR-1254-4

R.T.: 7.336 min
Delta R.T.: -0.004 min
Response: 4014740
Conc: 62.53 ng/ml



#29 AR-1254-4

R.T.: 6.437 min
Delta R.T.: -0.002 min
Response: 5457501
Conc: 49.50 ng/ml




#30 AR-1254-5

R.T.: 7.757 min
Delta R.T.: -0.003 min
Response: 4835280
Conc: 65.96 ng/ml

Instrument :
ECD_P
ClientSampleId :
02-A-02-B-02-C

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/13/2023
Supervised By :Ankita Jodhani 12/13/2023



#30 AR-1254-5

R.T.: 6.859 min
Delta R.T.: -0.003 min
Response: 21291313
Conc: 114.74 ng/ml