

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121223\
 Data File : PP062434.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Dec 2023 00:55
 Operator : YP\AJ
 Sample : 05760-05
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 SLAB

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 13 02:00:01 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Dec 12 05:13:02 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.432	3.718	34233068	51806295	20.407	16.882
2) SA Decachlor...	10.204	8.832	29006961	57049480	27.758	24.772

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121223\
Data File : PP062434.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Dec 2023 00:55
Operator : YP\AJ
Sample : 05760-05
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
SLAB

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 13 02:00:01 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121123.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Dec 12 05:13:02 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.432 min
Delta R.T.: -0.002 min
Response: 34233068
Conc: 20.41 ng/ml

Instrument :
ECD_P
ClientSampleId :
SLAB




#1 Tetrachloro-m-xylene

R.T.: 3.718 min
Delta R.T.: -0.002 min
Response: 51806295
Conc: 16.88 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.204 min
Delta R.T.: -0.009 min
Response: 29006961
Conc: 27.76 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.832 min
Delta R.T.: -0.006 min
Response: 57049480
Conc: 24.77 ng/ml