

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121225\
 Data File : PP077003.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Dec 2025 04:24
 Operator : YP\AJ
 Sample : Q3530-09
 Misc : AR1268 MDL 25PPB WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 MDL-WATER-03-QT4-2025

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 12/15/2025
 Supervised By :Yogesh Patel 12/15/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 15 00:28:42 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Nov 21 22:52:01 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.613	3.770	42901209	42585847	23.373	20.445
2) SA Decachlor...	10.360	8.758	37879588	36643245	25.275	20.767
Target Compounds						
41) L9 AR-1268-1	8.832	7.651	5815620	16100715	27.060	57.229 #
42) L9 AR-1268-2	8.927	7.714	5242328	12589437	26.051	49.462 #
43) L9 AR-1268-3	9.163	7.919	4490075	4547529	27.024	22.030
44) L9 AR-1268-4	9.588	8.208	2042822	1745211	24.524	18.582
45) L9 AR-1268-5	10.012	8.502	13665786	12896305	27.097	20.260 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121225\
Data File : PP077003.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Dec 2025 04:24
Operator : YP\AJ
Sample : Q3530-09
Misc : AR1268 MDL 25PPB WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
MDL-WATER-03-QT4-2025

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 12/15/2025
Supervised By :Yogesh Patel 12/15/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 15 00:28:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP112125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Nov 21 22:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

