

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121318\
 Data File : PP028438.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 13 Dec 2018 9:34
 Operator : EI\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 13 10:09:51 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP120718.M
 Quant Title : 8080.M
 QLast Update : Fri Dec 07 14:01:32 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	12.096	12.208	198.4E6	343.4E6	759.894	766.784
Target Compounds							
1) T	Dalapon	3.343	3.012	372.5E6	362.5E6	637.304	655.946
2) T	3,5-DICHL...	10.508	10.299	348.7E6	386.7E6	700.780	714.221
3) T	4-Nitroph...	11.673	11.352	304.0E6	162.9E6	698.207	728.068
5) T	DICAMBA	12.417	12.518	1143.6E6	1252.9E6	703.886	717.463
6) T	MCPD	12.760	12.710	34814006	68745009	76.113	77.659
7) T	MCPA	13.004	13.100	54485571	104.1E6	74.169	77.062
8) T	DICHLORPROP	13.615	13.633	254.8E6	310.1E6	531.318	716.360 #
9) T	2,4-D	13.966	14.120	271.6E6	368.1E6	727.405	724.984
10) T	Pentachlo...	14.384	14.732	5760.4E6	4802.9E6	711.787	719.174
11) T	2,4,5-TP ...	15.247	15.302	2098.0E6	1861.6E6	714.294	728.284
12) T	2,4,5-T	15.637	15.943	1893.6E6	1556.9E6	711.767	716.485
13) T	2,4-DB	16.400	16.544	158.8E6	201.2E6	743.376	750.595
14) T	DINOSEB	17.889	16.937	1488.5E6	1270.6E6	701.035	724.927
15) T	Picloram	17.652	18.317	2896.2E6	2476.8E6	696.783	707.523
16) T	DCPA	18.253	18.198	3117.9E6	2417.3E6	726.376	738.606

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121318\
Data File : PP028438.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 13 Dec 2018 9:34
Operator : EI\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 13 10:09:51 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP120718.M
Quant Title : 8080.M
QLast Update : Fri Dec 07 14:01:32 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

