

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121321\
 Data File : PP041955.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 13 Dec 2021 21:55
 Operator : AJ\MA
 Sample : AR1254ICC750
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 14 02:51:45 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121321.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Dec 14 02:43:53 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.929	4.071	910009	2018364	75.008	75.514
2) SA Decachlor...	10.925	9.438	838318	1387616	75.383	74.683
Target Compounds						
26) L6 AR-1254-1	7.148	6.224	284817	991789	751.342	749.371
27) L6 AR-1254-2	7.378	6.383	421075	849151	748.611	748.945
28) L6 AR-1254-3	7.757	6.808	452153	1337117	754.350	749.694
29) L6 AR-1254-4	8.053	7.047	322435	810060	736.962	748.953
30) L6 AR-1254-5	8.483	7.481	365352	1078028	741.366	751.126



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121321\
Data File : PP041955.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 13 Dec 2021 21:55
Operator : AJ\MA
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 14 02:51:45 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121321.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Dec 14 02:43:53 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.929 min
Delta R.T.: 0.000 min
Response: 910009
Conc: 75.01 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750




#1 Tetrachloro-m-xylene

R.T.: 4.071 min
Delta R.T.: 0.000 min
Response: 2018364
Conc: 75.51 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.925 min
Delta R.T.: 0.000 min
Response: 838318
Conc: 75.38 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.438 min
Delta R.T.: 0.000 min
Response: 1387616
Conc: 74.68 ng/ml



#26 AR-1254-1


R.T.: 7.148 min
Delta R.T.: 0.000 min
Response: 284817
Conc: 751.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254\ICC750




#26 AR-1254-1

R.T.: 6.224 min
Delta R.T.: 0.000 min
Response: 991789
Conc: 749.37 ng/ml




#27 AR-1254-2

R.T.: 7.378 min
Delta R.T.: 0.000 min
Response: 421075
Conc: 748.61 ng/ml



#27 AR-1254-2


R.T.: 6.383 min
Delta R.T.: 0.000 min
Response: 849151
Conc: 748.94 ng/ml



#28 AR-1254-3


R.T.: 7.757 min
Delta R.T.: 0.000 min
Response: 452153
Conc: 754.35 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254\ICC750




#28 AR-1254-3

R.T.: 6.808 min
Delta R.T.: 0.000 min
Response: 1337117
Conc: 749.69 ng/ml




#29 AR-1254-4

R.T.: 8.053 min
Delta R.T.: 0.000 min
Response: 322435
Conc: 736.96 ng/ml



#29 AR-1254-4


R.T.: 7.047 min
Delta R.T.: 0.000 min
Response: 810060
Conc: 748.95 ng/ml



#30 AR-1254-5

R.T.: 8.483 min
Delta R.T.: 0.000 min
Response: 365352
Conc: 741.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750



#30 AR-1254-5

R.T.: 7.481 min
Delta R.T.: 0.000 min
Response: 1078028
Conc: 751.13 ng/ml