

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121321\
 Data File : PP041956.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 13 Dec 2021 22:12
 Operator : AJ\MA
 Sample : AR1254ICC500
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 14 02:46:27 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121321.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Dec 14 02:43:53 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.929	4.071	638896	1379292	50.000	50.000
2) SA Decachlor...	10.922	9.439	575048	981561	50.000	50.000
Target Compounds						
26) L6 AR-1254-1	7.148	6.224	201326	699881	500.000	500.000
27) L6 AR-1254-2	7.378	6.383	300593	600154	500.000	500.000
28) L6 AR-1254-3	7.757	6.809	316316	935861	500.000	500.000
29) L6 AR-1254-4	8.052	7.048	239377	569789	500.000	500.000
30) L6 AR-1254-5	8.483	7.482	261077	753626	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121321\
Data File : PP041956.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 13 Dec 2021 22:12
Operator : AJ\MA
Sample : AR1254ICC500
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 14 02:46:27 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121321.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Dec 14 02:43:53 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.929 min
Delta R.T.: 0.000 min
Response: 638896
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500




#1 Tetrachloro-m-xylene

R.T.: 4.071 min
Delta R.T.: 0.000 min
Response: 1379292
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.922 min
Delta R.T.: 0.000 min
Response: 575048
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.439 min
Delta R.T.: 0.000 min
Response: 981561
Conc: 50.00 ng/ml



#26 AR-1254-1


R.T.: 7.148 min
Delta R.T.: 0.000 min
Response: 201326
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254\ICC500




#26 AR-1254-1

R.T.: 6.224 min
Delta R.T.: 0.000 min
Response: 699881
Conc: 500.00 ng/ml




#27 AR-1254-2

R.T.: 7.378 min
Delta R.T.: 0.000 min
Response: 300593
Conc: 500.00 ng/ml



#27 AR-1254-2


R.T.: 6.383 min
Delta R.T.: 0.000 min
Response: 600154
Conc: 500.00 ng/ml



#28 AR-1254-3


R.T.: 7.757 min
Delta R.T.: 0.000 min
Response: 316316
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254\ICC500




#28 AR-1254-3

R.T.: 6.809 min
Delta R.T.: 0.000 min
Response: 935861
Conc: 500.00 ng/ml




#29 AR-1254-4

R.T.: 8.052 min
Delta R.T.: 0.000 min
Response: 239377
Conc: 500.00 ng/ml



#29 AR-1254-4


R.T.: 7.048 min
Delta R.T.: 0.000 min
Response: 569789
Conc: 500.00 ng/ml



#30 AR-1254-5

R.T.: 8.483 min
Delta R.T.: 0.000 min
Response: 261077
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254\ICC500



#30 AR-1254-5

R.T.: 7.482 min
Delta R.T.: 0.000 min
Response: 753626
Conc: 500.00 ng/ml