

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121321\
 Data File : PP041964.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 14 Dec 2021 00:25
 Operator : AJ\MA
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP121321

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 14 03:00:00 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121321.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Dec 14 02:43:53 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.929	4.070	644033	1357502	52.631	51.464
2) SA Decachlor...	10.922	9.436	576880	986762	53.495	51.516
Target Compounds						
26) L6 AR-1254-1	7.147	6.223	206379	698461	518.673	508.628
27) L6 AR-1254-2	7.377	6.382	307958	604249	522.604	513.115
28) L6 AR-1254-3	7.756	6.807	323911	937344	521.418	511.633
29) L6 AR-1254-4	8.052	7.046	230509	572431	503.370	521.655
30) L6 AR-1254-5	8.482	7.480	269997	756133	527.256	524.372



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121321\
Data File : PP041964.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 14 Dec 2021 00:25
Operator : AJ\MA
Sample : AR1254ICV500
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP121321

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 14 03:00:00 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121321.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Dec 14 02:43:53 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.929 min
Delta R.T.: 0.000 min
Response: 644033
Conc: 52.63 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP121321




#1 Tetrachloro-m-xylene

R.T.: 4.070 min
Delta R.T.: 0.000 min
Response: 1357502
Conc: 51.46 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.922 min
Delta R.T.: 0.001 min
Response: 576880
Conc: 53.50 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.436 min
Delta R.T.: -0.002 min
Response: 986762
Conc: 51.52 ng/ml



#26 AR-1254-1


R.T.: 7.147 min
Delta R.T.: 0.000 min
Response: 206379
Conc: 518.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP121321




#26 AR-1254-1

R.T.: 6.223 min
Delta R.T.: 0.000 min
Response: 698461
Conc: 508.63 ng/ml




#27 AR-1254-2

R.T.: 7.377 min
Delta R.T.: 0.000 min
Response: 307958
Conc: 522.60 ng/ml



#27 AR-1254-2


R.T.: 6.382 min
Delta R.T.: 0.000 min
Response: 604249
Conc: 513.11 ng/ml



#28 AR-1254-3


R.T.: 7.756 min
Delta R.T.: 0.002 min
Response: 323911
Conc: 521.42 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP121321




#28 AR-1254-3

R.T.: 6.807 min
Delta R.T.: -0.001 min
Response: 937344
Conc: 511.63 ng/ml




#29 AR-1254-4

R.T.: 8.052 min
Delta R.T.: -0.001 min
Response: 230509
Conc: 503.37 ng/ml



#29 AR-1254-4


R.T.: 7.046 min
Delta R.T.: -0.001 min
Response: 572431
Conc: 521.66 ng/ml



#30 AR-1254-5

R.T.: 8.482 min
Delta R.T.: -0.003 min
Response: 269997
Conc: 527.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP121321



#30 AR-1254-5

R.T.: 7.480 min
Delta R.T.: -0.001 min
Response: 756133
Conc: 524.37 ng/ml