

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121322\
 Data File : PP053725.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Dec 2022 00:27
 Operator : YP\AJ
 Sample : AR1268ICC1000
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 14 04:59:00 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121322.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Dec 14 04:58:26 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.339	3.592	214.8E6	163.2E6	95.775	94.854
2) SA Decachlor...	10.102	8.636	267.2E6	226.8E6	92.290	93.423
Target Compounds						
41) L9 AR-1268-1	8.595	7.526	227.6E6	172.9E6	934.820	935.587
42) L9 AR-1268-2	8.689	7.589	211.9E6	167.3E6	923.286	951.346
43) L9 AR-1268-3	8.918	7.799	178.2E6	143.0E6	923.829	944.479
44) L9 AR-1268-4	9.344	8.085	71311413	56459098	907.027	936.744
45) L9 AR-1268-5	9.761	8.378	567.1E6	468.7E6	948.448	961.393



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121322\
Data File : PP053725.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Dec 2022 00:27
Operator : YP\AJ
Sample : AR1268ICC1000
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 14 04:59:00 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121322.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Dec 14 04:58:26 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.339 min
Delta R.T.: 0.001 min
Response: 214785348
Conc: 95.78 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.592 min
Delta R.T.: 0.000 min
Response: 163235469
Conc: 94.85 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.102 min
Delta R.T.: 0.000 min
Response: 267180025
Conc: 92.29 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.636 min
Delta R.T.: 0.001 min
Response: 226816901
Conc: 93.42 ng/ml



#41 AR-1268-1


R.T.: 8.595 min
Delta R.T.: 0.001 min
Response: 227618737
Conc: 934.82 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.526 min
Delta R.T.: 0.001 min
Response: 172938233
Conc: 935.59 ng/ml




#42 AR-1268-2

R.T.: 8.689 min
Delta R.T.: 0.002 min
Response: 211932048
Conc: 923.29 ng/ml



#42 AR-1268-2


R.T.: 7.589 min
Delta R.T.: 0.000 min
Response: 167317977
Conc: 951.35 ng/ml



#43 AR-1268-3


R.T.: 8.918 min
Delta R.T.: 0.000 min
Response: 178216112
Conc: 923.83 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.799 min
Delta R.T.: 0.000 min
Response: 143022571
Conc: 944.48 ng/ml




#44 AR-1268-4

R.T.: 9.344 min
Delta R.T.: 0.000 min
Response: 71311413
Conc: 907.03 ng/ml



#44 AR-1268-4


R.T.: 8.085 min
Delta R.T.: 0.000 min
Response: 56459098
Conc: 936.74 ng/ml



#45 AR-1268-5

R.T.: 9.761 min
Delta R.T.: 0.000 min
Response: 567075122
Conc: 948.45 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.378 min
Delta R.T.: 0.001 min
Response: 468741390
Conc: 961.39 ng/ml