

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121322\
 Data File : PP053704.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Dec 2022 18:43
 Operator : YP\AJ
 Sample : AR1660ICC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 14 01:26:50 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121322.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Dec 14 01:11:02 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.339	3.594	106.8E6	80834379	50.000	50.000
2) SA Decachlor...	10.102	8.636	84694942	70486629	50.000	50.000
Target Compounds						
3) L1 AR-1016-1	5.512	4.686	34303554	22969465	500.000	500.000
4) L1 AR-1016-2	5.534	4.705	50206507	31834513	500.000	500.000
5) L1 AR-1016-3	5.596	4.882	31886371	17240144	500.000	500.000
6) L1 AR-1016-4	5.695	4.924	25264608	15017260	500.000	500.000
7) L1 AR-1016-5	5.991	5.139	26594548	18348052	500.000	500.000
31) L7 AR-1260-1	7.123	6.179	47939886	32849538	500.000	500.000
32) L7 AR-1260-2	7.381	6.367	54760758	37963487	500.000	500.000
33) L7 AR-1260-3	7.742	6.522	38059948	36485810	500.000	500.000
34) L7 AR-1260-4	7.968	6.997	46250878	27712093	500.000	500.000
35) L7 AR-1260-5	8.285	7.239	78606430	60759568	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121322\
Data File : PP053704.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Dec 2022 18:43
Operator : YP\AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 14 01:26:50 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121322.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Dec 14 01:11:02 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.339 min
Delta R.T.: 0.000 min
Response: 106788600
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.594 min
Delta R.T.: 0.000 min
Response: 80834379
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.102 min
Delta R.T.: 0.000 min
Response: 84694942
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.636 min
Delta R.T.: 0.000 min
Response: 70486629
Conc: 50.00 ng/ml



#3 AR-1016-1


R.T.: 5.512 min
Delta R.T.: 0.000 min
Response: 34303554
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500




#3 AR-1016-1

R.T.: 4.686 min
Delta R.T.: 0.000 min
Response: 22969465
Conc: 500.00 ng/ml




#4 AR-1016-2

R.T.: 5.534 min
Delta R.T.: 0.000 min
Response: 50206507
Conc: 500.00 ng/ml



#4 AR-1016-2


R.T.: 4.705 min
Delta R.T.: 0.000 min
Response: 31834513
Conc: 500.00 ng/ml



#5 AR-1016-3


R.T.: 5.596 min
Delta R.T.: 0.000 min
Response: 31886371
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500




#5 AR-1016-3

R.T.: 4.882 min
Delta R.T.: 0.000 min
Response: 17240144
Conc: 500.00 ng/ml




#6 AR-1016-4

R.T.: 5.695 min
Delta R.T.: 0.000 min
Response: 25264608
Conc: 500.00 ng/ml



#6 AR-1016-4


R.T.: 4.924 min
Delta R.T.: 0.000 min
Response: 15017260
Conc: 500.00 ng/ml



#7 AR-1016-5


R.T.: 5.991 min
Delta R.T.: 0.000 min
Response: 26594548
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500




#7 AR-1016-5

R.T.: 5.139 min
Delta R.T.: 0.000 min
Response: 18348052
Conc: 500.00 ng/ml




#31 AR-1260-1

R.T.: 7.123 min
Delta R.T.: 0.000 min
Response: 47939886
Conc: 500.00 ng/ml



#31 AR-1260-1


R.T.: 6.179 min
Delta R.T.: 0.000 min
Response: 32849538
Conc: 500.00 ng/ml



#32 AR-1260-2


R.T.: 7.381 min
Delta R.T.: 0.000 min
Response: 54760758
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500




#32 AR-1260-2

R.T.: 6.367 min
Delta R.T.: 0.000 min
Response: 37963487
Conc: 500.00 ng/ml




#33 AR-1260-3

R.T.: 7.742 min
Delta R.T.: 0.000 min
Response: 38059948
Conc: 500.00 ng/ml



#33 AR-1260-3


R.T.: 6.522 min
Delta R.T.: 0.000 min
Response: 36485810
Conc: 500.00 ng/ml



#34 AR-1260-4


R.T.: 7.968 min
Delta R.T.: 0.000 min
Response: 46250878
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500




#34 AR-1260-4

R.T.: 6.997 min
Delta R.T.: 0.000 min
Response: 27712093
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 8.285 min
Delta R.T.: 0.000 min
Response: 78606430
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 7.239 min
Delta R.T.: 0.000 min
Response: 60759568
Conc: 500.00 ng/ml