

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121322\
 Data File : PP053706.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Dec 2022 19:16
 Operator : YP\AJ
 Sample : AR1660ICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 14 02:39:32 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121322.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Dec 14 02:39:24 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.340	3.594	10072293	7098400	4.797	4.545
2) SA Decachlor...	10.103	8.636	9069445	7083458	5.342	5.071
Target Compounds						
3) L1 AR-1016-1	5.513	4.686	3309754	2279176	48.810	49.627
4) L1 AR-1016-2	5.536	4.706	5023538	2978387	50.432	47.534
5) L1 AR-1016-3	5.598	4.883	3193665	1650814	50.532	48.637
6) L1 AR-1016-4	5.697	4.925	2391051	1523632	48.253	51.131
7) L1 AR-1016-5	5.992	5.139	2644334	1812448	50.238	49.983
31) L7 AR-1260-1	7.123	6.179	5205015	3307466	53.897	50.705
32) L7 AR-1260-2	7.381	6.368	5947550	3775118	53.838	50.131
33) L7 AR-1260-3	7.742	6.523	3951008	3673342	52.255	50.678
34) L7 AR-1260-4	7.968	6.998	5208863	2631983	55.499	48.508
35) L7 AR-1260-5	8.285	7.240	8484763	5864140	53.508	49.037



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121322\
Data File : PP053706.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Dec 2022 19:16
Operator : YP\AJ
Sample : AR1660ICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 14 02:39:32 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121322.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Dec 14 02:39:24 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.340 min
Delta R.T.: 0.000 min
Response: 10072293
Conc: 4.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050




#1 Tetrachloro-m-xylene

R.T.: 3.594 min
Delta R.T.: 0.000 min
Response: 7098400
Conc: 4.55 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.103 min
Delta R.T.: 0.000 min
Response: 9069445
Conc: 5.34 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.636 min
Delta R.T.: 0.000 min
Response: 7083458
Conc: 5.07 ng/ml



#3 AR-1016-1


R.T.: 5.513 min
Delta R.T.: 0.000 min
Response: 3309754
Conc: 48.81 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050




#3 AR-1016-1

R.T.: 4.686 min
Delta R.T.: 0.000 min
Response: 2279176
Conc: 49.63 ng/ml




#4 AR-1016-2

R.T.: 5.536 min
Delta R.T.: 0.000 min
Response: 5023538
Conc: 50.43 ng/ml



#4 AR-1016-2


R.T.: 4.706 min
Delta R.T.: 0.000 min
Response: 2978387
Conc: 47.53 ng/ml



#5 AR-1016-3


R.T.: 5.598 min
Delta R.T.: 0.000 min
Response: 3193665
Conc: 50.53 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050




#5 AR-1016-3

R.T.: 4.883 min
Delta R.T.: 0.000 min
Response: 1650814
Conc: 48.64 ng/ml




#6 AR-1016-4

R.T.: 5.697 min
Delta R.T.: 0.000 min
Response: 2391051
Conc: 48.25 ng/ml



#6 AR-1016-4


R.T.: 4.925 min
Delta R.T.: 0.000 min
Response: 1523632
Conc: 51.13 ng/ml



#7 AR-1016-5


R.T.: 5.992 min
Delta R.T.: 0.000 min
Response: 2644334
Conc: 50.24 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050




#7 AR-1016-5

R.T.: 5.139 min
Delta R.T.: 0.000 min
Response: 1812448
Conc: 49.98 ng/ml




#31 AR-1260-1

R.T.: 7.123 min
Delta R.T.: 0.000 min
Response: 5205015
Conc: 53.90 ng/ml



#31 AR-1260-1


R.T.: 6.179 min
Delta R.T.: 0.000 min
Response: 3307466
Conc: 50.71 ng/ml



#32 AR-1260-2


R.T.: 7.381 min
Delta R.T.: 0.000 min
Response: 5947550
Conc: 53.84 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050




#32 AR-1260-2

R.T.: 6.368 min
Delta R.T.: 0.000 min
Response: 3775118
Conc: 50.13 ng/ml




#33 AR-1260-3

R.T.: 7.742 min
Delta R.T.: 0.000 min
Response: 3951008
Conc: 52.26 ng/ml



#33 AR-1260-3


R.T.: 6.523 min
Delta R.T.: 0.000 min
Response: 3673342
Conc: 50.68 ng/ml



#34 AR-1260-4


R.T.: 7.968 min
Delta R.T.: 0.000 min
Response: 5208863
Conc: 55.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050




#34 AR-1260-4

R.T.: 6.998 min
Delta R.T.: 0.000 min
Response: 2631983
Conc: 48.51 ng/ml



#35 AR-1260-5

R.T.: 8.285 min
Delta R.T.: 0.000 min
Response: 8484763
Conc: 53.51 ng/ml



#35 AR-1260-5

R.T.: 7.240 min
Delta R.T.: 0.000 min
Response: 5864140
Conc: 49.04 ng/ml