

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121422\
 Data File : PP053763.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Dec 2022 18:53
 Operator : YP\AJ
 Sample : N6046-03
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 HR-02-121322

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/15/2022
 Supervised By :Ankita Jodhani 12/15/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 14 21:47:48 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121322.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Dec 14 05:38:40 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.339	3.592	42831494	33367921	20.398	21.366
2) SA Decachlor...	10.101	8.633	33822322	25970587	19.921	18.594

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121422\
Data File : PP053763.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Dec 2022 18:53
Operator : YP\AJ
Sample : N6046-03
Misc :
ALS Vial : 24 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
HR-02-121322


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/15/2022
Supervised By :Ankita Jodhani 12/15/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 14 21:47:48 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121322.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Dec 14 05:38:40 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.339 min
Delta R.T.: 0.000 min
Response: 42831494
Conc: 20.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
HR-02-121322


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/15/2022
Supervised By :Ankita Jodhani 12/15/2022




#1 Tetrachloro-m-xylene

R.T.: 3.592 min
Delta R.T.: -0.002 min
Response: 33367921
Conc: 21.37 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.101 min
Delta R.T.: -0.001 min
Response: 33822322
Conc: 19.92 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.633 min
Delta R.T.: -0.003 min
Response: 25970587
Conc: 18.59 ng/ml