

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121522\
 Data File : PP053786.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Dec 2022 16:35
 Operator : YP\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 15 21:16:08 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121322.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Dec 14 05:38:40 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.339	3.592	48558239	35442014	23.126	22.694
2)	SA Decachlor...	10.102	8.631	37887638	31621460	22.316	22.639
Target Compounds							
3)	L1 AR-1016-1	5.506	4.699	75835	319464	1.118	6.956 #
4)	L1 AR-1016-2	5.536	4.699	140681	319464	1.412	5.099 #
5)	L1 AR-1016-3	5.584	0.000	503921	0	7.973	N.D. #
6)	L1 AR-1016-4	5.699	4.939	12365	126569	0.250	4.247 #
8)	L2 AR-1221-1	4.553	0.000	17420	0	0.676	N.D. #
9)	L2 AR-1221-2	4.646	3.892	747074	467801	38.373	32.003
10)	L2 AR-1221-3	4.710	3.963	123746	47508	2.129	1.068 #
11)	L3 AR-1232-1	4.710	3.963	123746	47508	2.763	1.381 #
12)	L3 AR-1232-2	5.257f	4.699	61655	319464	2.397	10.704 #
13)	L3 AR-1232-3	5.536	0.000	140681	0	2.997	N.D. #
14)	L3 AR-1232-4	5.699	4.978	12365	318170	0.522	20.590 #
16)	L4 AR-1242-1	5.506	4.699	75835	319464	1.448	9.021 #
17)	L4 AR-1242-2	5.536	4.699	140681	319464	1.846	6.650 #
18)	L4 AR-1242-3	5.584	0.000	503921	0	10.285	N.D. #
19)	L4 AR-1242-4	5.699	4.978	12365	318170	0.322	11.153 #
20)	L4 AR-1242-5	6.426	5.516f	412864	857079	10.123	28.080 #
21)	L5 AR-1248-1	5.506	4.699	75835	319464	1.799	11.230 #
22)	L5 AR-1248-2	5.784	4.939	146716	126569	2.321	3.060 #
23)	L5 AR-1248-3	5.993	4.978	214805	318170	3.067	7.576 #
25)	L5 AR-1248-5	6.426	5.516f	412864	857079	5.859	22.826 #
26)	L6 AR-1254-1	6.371	5.516f	1522811	857079	19.651	13.089 #
27)	L6 AR-1254-2	6.592	5.652	541852	991469	4.657	17.001 #
28)	L6 AR-1254-3	6.957	6.054	683129	418230	5.832	4.754
29)	L6 AR-1254-4	7.246	6.273	1025124	232270	12.931	5.070 #
30)	L6 AR-1254-5	7.672	6.695	143564	231628	1.472	2.972 #
31)	L7 AR-1260-1	7.119	6.182	129745	354100	1.343	5.429 #
32)	L7 AR-1260-2	7.397f	6.346f	461392	848036	4.177	11.261 #
33)	L7 AR-1260-3	7.735	6.527	87529	105793	1.158	1.460 #
34)	L7 AR-1260-4	7.994f	6.997	1190796	84289	12.688	1.553 #
35)	L7 AR-1260-5	8.286	7.240	291919	254011	1.841	2.124
36)	L8 AR-1262-1	7.735	6.762f	87529	28151	0.731	0.789
37)	L8 AR-1262-2	8.286	6.997	291919	84289	1.522	1.084 #
38)	L8 AR-1262-3	8.593	7.515	244038	335906	1.761	5.331 #
39)	L8 AR-1262-4	8.688	0.000	265930	0	4.273	N.D. #
40)	L8 AR-1262-5	9.359	8.079	203738	301257	2.883	6.047 #
41)	L9 AR-1268-1	8.593	7.515	244038	335906	1.000	1.857 #
42)	L9 AR-1268-2	8.688	0.000	265930	0	1.170	N.D. #
43)	L9 AR-1268-3	8.923	7.800	106610	119634	0.556	0.810 #
44)	L9 AR-1268-4	9.359f	8.079	203738	301257	2.647	5.406 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121522\
 Data File : PP053786.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Dec 2022 16:35
 Operator : YP\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 15 21:16:08 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121322.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Dec 14 05:38:40 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
45) L9	AR-1268-5	9.759	8.375	106069	238724	0.180	0.505 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121522\
Data File : PP053786.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Dec 2022 16:35
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 15 21:16:08 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121322.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Dec 14 05:38:40 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.339 min
Delta R.T.: 0.000 min
Response: 48558239
Conc: 23.13 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.592 min
Delta R.T.: -0.001 min
Response: 35442014
Conc: 22.69 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.102 min
Delta R.T.: 0.000 min
Response: 37887638
Conc: 22.32 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.631 min
Delta R.T.: -0.004 min
Response: 31621460
Conc: 22.64 ng/ml



#3 AR-1016-1


R.T.: 5.506 min
Delta R.T.: -0.005 min
Response: 75835
Conc: 1.12 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.699 min
Delta R.T.: 0.013 min
Response: 319464
Conc: 6.96 ng/ml




#4 AR-1016-2

R.T.: 5.536 min
Delta R.T.: 0.001 min
Response: 140681
Conc: 1.41 ng/ml



#4 AR-1016-2


R.T.: 4.699 min
Delta R.T.: -0.006 min
Response: 319464
Conc: 5.10 ng/ml



#5 AR-1016-3


R.T.: 5.584 min
Delta R.T.: -0.013 min
Response: 503921
Conc: 7.97 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 0.000 min
Exp R.T. : 4.882 min
Response: 0
Conc: N.D.




#6 AR-1016-4

R.T.: 5.699 min
Delta R.T.: 0.003 min
Response: 12365
Conc: 0.25 ng/ml



#6 AR-1016-4


R.T.: 4.939 min
Delta R.T.: 0.014 min
Response: 126569
Conc: 4.25 ng/ml



#7 AR-1016-5


R.T.: 5.993 min
Delta R.T.: 0.002 min
Response: 214805
Conc: 4.08 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.155 min
Delta R.T.: 0.016 min
Response: -276051
Conc: N.D.




#8 AR-1221-1

R.T.: 4.553 min
Delta R.T.: 0.004 min
Response: 17420
Conc: 0.68 ng/ml



#8 AR-1221-1


R.T.: 0.000 min
Exp R.T. : 3.810 min
Response: 0
Conc: N.D.



#9 AR-1221-2


R.T.: 4.646 min
Delta R.T.: 0.011 min
Response: 747074
Conc: 38.37 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.892 min
Delta R.T.: -0.003 min
Response: 467801
Conc: 32.00 ng/ml




#10 AR-1221-3

R.T.: 4.710 min
Delta R.T.: -0.002 min
Response: 123746
Conc: 2.13 ng/ml



#10 AR-1221-3


R.T.: 3.963 min
Delta R.T.: -0.008 min
Response: 47508
Conc: 1.07 ng/ml



#11 AR-1232-1


R.T.: 4.710 min
Delta R.T.: -0.001 min
Response: 123746
Conc: 2.76 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 3.963 min
Delta R.T.: -0.008 min
Response: 47508
Conc: 1.38 ng/ml




#12 AR-1232-2

R.T.: 5.257 min
Delta R.T.: 0.015 min
Response: 61655
Conc: 2.40 ng/ml



#12 AR-1232-2


R.T.: 4.699 min
Delta R.T.: -0.006 min
Response: 319464
Conc: 10.70 ng/ml



#13 AR-1232-3


R.T.: 5.536 min
Delta R.T.: 0.002 min
Response: 140681
Conc: 3.00 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 0.000 min
Exp R.T. : 4.882 min
Response: 0
Conc: N.D.




#14 AR-1232-4

R.T.: 5.699 min
Delta R.T.: 0.004 min
Response: 12365
Conc: 0.52 ng/ml



#14 AR-1232-4


R.T.: 4.978 min
Delta R.T.: 0.012 min
Response: 318170
Conc: 20.59 ng/ml



#15 AR-1232-5


R.T.: 5.784 min
Delta R.T.: -0.002 min
Response: 146716
Conc: 7.96 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.155 min
Delta R.T.: 0.017 min
Response: -276051
Conc: N.D.




#16 AR-1242-1

R.T.: 5.506 min
Delta R.T.: -0.005 min
Response: 75835
Conc: 1.45 ng/ml



#16 AR-1242-1


R.T.: 4.699 min
Delta R.T.: 0.014 min
Response: 319464
Conc: 9.02 ng/ml



#17 AR-1242-2


R.T.: 5.536 min
Delta R.T.: 0.001 min
Response: 140681
Conc: 1.85 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.699 min
Delta R.T.: -0.005 min
Response: 319464
Conc: 6.65 ng/ml




#18 AR-1242-3

R.T.: 5.584 min
Delta R.T.: -0.013 min
Response: 503921
Conc: 10.29 ng/ml



#18 AR-1242-3


R.T.: 0.000 min
Exp R.T.: 4.882 min
Response: 0
Conc: N.D.



#19 AR-1242-4


R.T.: 5.699 min
Delta R.T.: 0.004 min
Response: 12365
Conc: 0.32 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 4.978 min
Delta R.T.: 0.012 min
Response: 318170
Conc: 11.15 ng/ml




#20 AR-1242-5

R.T.: 6.426 min
Delta R.T.: -0.010 min
Response: 412864
Conc: 10.12 ng/ml



#20 AR-1242-5


R.T.: 5.516 min
Delta R.T.: 0.024 min
Response: 857079
Conc: 28.08 ng/ml



#21 AR-1248-1


R.T.: 5.506 min
Delta R.T.: -0.005 min
Response: 75835
Conc: 1.80 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.699 min
Delta R.T.: 0.013 min
Response: 319464
Conc: 11.23 ng/ml




#22 AR-1248-2

R.T.: 5.784 min
Delta R.T.: -0.002 min
Response: 146716
Conc: 2.32 ng/ml



#22 AR-1248-2


R.T.: 4.939 min
Delta R.T.: 0.014 min
Response: 126569
Conc: 3.06 ng/ml



#23 AR-1248-3


R.T.: 5.993 min
Delta R.T.: 0.002 min
Response: 214805
Conc: 3.07 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 4.978 min
Delta R.T.: 0.012 min
Response: 318170
Conc: 7.58 ng/ml




#24 AR-1248-4

R.T.: 6.408 min
Delta R.T.: 0.011 min
Response: 273288
Conc: 3.87 ng/ml



#24 AR-1248-4


R.T.: 5.155 min
Delta R.T.: 0.017 min
Response: -276051
Conc: N.D.



#25 AR-1248-5


R.T.: 6.426 min
Delta R.T.: -0.010 min
Response: 412864
Conc: 5.86 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.516 min
Delta R.T.: -0.017 min
Response: 857079
Conc: 22.83 ng/ml




#26 AR-1254-1

R.T.: 6.371 min
Delta R.T.: 0.002 min
Response: 1522811
Conc: 19.65 ng/ml



#26 AR-1254-1


R.T.: 5.516 min
Delta R.T.: 0.024 min
Response: 857079
Conc: 13.09 ng/ml



#27 AR-1254-2


R.T.: 6.592 min
Delta R.T.: 0.002 min
Response: 541852
Conc: 4.66 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.652 min
Delta R.T.: 0.011 min
Response: 991469
Conc: 17.00 ng/ml




#28 AR-1254-3

R.T.: 6.957 min
Delta R.T.: 0.000 min
Response: 683129
Conc: 5.83 ng/ml



#28 AR-1254-3


R.T.: 6.054 min
Delta R.T.: 0.007 min
Response: 418230
Conc: 4.75 ng/ml



#29 AR-1254-4


R.T.: 7.246 min
Delta R.T.: 0.002 min
Response: 1025124
Conc: 12.93 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.273 min
Delta R.T.: -0.003 min
Response: 232270
Conc: 5.07 ng/ml




#30 AR-1254-5

R.T.: 7.672 min
Delta R.T.: 0.008 min
Response: 143564
Conc: 1.47 ng/ml



#30 AR-1254-5


R.T.: 6.695 min
Delta R.T.: 0.000 min
Response: 231628
Conc: 2.97 ng/ml



#31 AR-1260-1


R.T.: 7.119 min
Delta R.T.: -0.003 min
Response: 129745
Conc: 1.34 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.182 min
Delta R.T.: 0.003 min
Response: 354100
Conc: 5.43 ng/ml




#32 AR-1260-2

R.T.: 7.397 min
Delta R.T.: 0.016 min
Response: 461392
Conc: 4.18 ng/ml



#32 AR-1260-2


R.T.: 6.346 min
Delta R.T.: -0.021 min
Response: 848036
Conc: 11.26 ng/ml



#33 AR-1260-3


R.T.: 7.735 min
Delta R.T.: -0.006 min
Response: 87529
Conc: 1.16 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.527 min
Delta R.T.: 0.005 min
Response: 105793
Conc: 1.46 ng/ml




#34 AR-1260-4

R.T.: 7.994 min
Delta R.T.: 0.026 min
Response: 1190796
Conc: 12.69 ng/ml



#34 AR-1260-4


R.T.: 6.997 min
Delta R.T.: 0.000 min
Response: 84289
Conc: 1.55 ng/ml



#35 AR-1260-5


R.T.: 8.286 min
Delta R.T.: 0.002 min
Response: 291919
Conc: 1.84 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.240 min
Delta R.T.: 0.001 min
Response: 254011
Conc: 2.12 ng/ml




#36 AR-1262-1

R.T.: 7.735 min
Delta R.T.: -0.006 min
Response: 87529
Conc: 0.73 ng/ml



#36 AR-1262-1


R.T.: 6.762 min
Delta R.T.: -0.027 min
Response: 28151
Conc: 0.79 ng/ml



#37 AR-1262-2


R.T.: 8.286 min
Delta R.T.: 0.002 min
Response: 291919
Conc: 1.52 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 6.997 min
Delta R.T.: 0.000 min
Response: 84289
Conc: 1.08 ng/ml




#38 AR-1262-3

R.T.: 8.593 min
Delta R.T.: -0.005 min
Response: 244038
Conc: 1.76 ng/ml



#38 AR-1262-3


R.T.: 7.515 min
Delta R.T.: -0.010 min
Response: 335906
Conc: 5.33 ng/ml



#39 AR-1262-4


R.T.: 8.688 min
Delta R.T.: 0.001 min
Response: 265930
Conc: 4.27 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 0.000 min
Exp R.T. : 7.588 min
Response: 0
Conc: N.D.




#40 AR-1262-5

R.T.: 9.359 min
Delta R.T.: 0.015 min
Response: 203738
Conc: 2.88 ng/ml



#40 AR-1262-5


R.T.: 8.079 min
Delta R.T.: -0.005 min
Response: 301257
Conc: 6.05 ng/ml



#41 AR-1268-1


R.T.: 8.593 min
Delta R.T.: -0.002 min
Response: 244038
Conc: 1.00 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.515 min
Delta R.T.: -0.011 min
Response: 335906
Conc: 1.86 ng/ml




#42 AR-1268-2

R.T.: 8.688 min
Delta R.T.: -0.002 min
Response: 265930
Conc: 1.17 ng/ml



#42 AR-1268-2


R.T.: 0.000 min
Exp R.T. : 7.589 min
Response: 0
Conc: N.D.



#43 AR-1268-3


R.T.: 8.923 min
Delta R.T.: 0.005 min
Response: 106610
Conc: 0.56 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.800 min
Delta R.T.: 0.000 min
Response: 119634
Conc: 0.81 ng/ml




#44 AR-1268-4

R.T.: 9.359 min
Delta R.T.: 0.016 min
Response: 203738
Conc: 2.65 ng/ml



#44 AR-1268-4


R.T.: 8.079 min
Delta R.T.: -0.006 min
Response: 301257
Conc: 5.41 ng/ml



#45 AR-1268-5

R.T.: 9.759 min
Delta R.T.: -0.002 min
Response: 106069
Conc: 0.18 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.375 min
Delta R.T.: -0.003 min
Response: 238724
Conc: 0.50 ng/ml