

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121522\
 Data File : PP053811.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Dec 2022 23:46
 Operator : YP\AJ
 Sample : N6037-21
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 16 01:53:08 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121322.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Dec 14 05:38:40 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.340	3.592	43479618	31093977	20.707	19.910
2)	SA Decachlor...	10.107	8.633	23599231	19703692	13.900	14.107
Target Compounds							
3)	L1 AR-1016-1	5.509	4.697	110372	95853	1.628	2.087 #
4)	L1 AR-1016-2	5.536	4.697	151223	95853	1.518	1.530
5)	L1 AR-1016-3	5.590	0.000	237425	0	3.757	N.D. #
6)	L1 AR-1016-4	5.686	0.000	73242	0	1.478	N.D. #
7)	L1 AR-1016-5	5.982	5.127	135697	259784	2.578	7.164 #
8)	L2 AR-1221-1	4.547	3.815	1299396	147570	50.439	7.621 #
9)	L2 AR-1221-2	4.647	3.892	649121	1008964	33.342	69.025 #
10)	L2 AR-1221-3	4.701	0.000	38307	0	0.659	N.D. #
11)	L3 AR-1232-1	4.701	0.000	38307	0	0.855	N.D. #
12)	L3 AR-1232-2	5.251	4.697	77333	95853	3.006	3.212
13)	L3 AR-1232-3	5.536	0.000	151223	0	3.222	N.D. #
14)	L3 AR-1232-4	5.686	0.000	73242	0	3.091	N.D. #
15)	L3 AR-1232-5	5.787	5.127	68076	259784	3.691	15.776 #
16)	L4 AR-1242-1	5.509	4.697	110372	95853	2.107	2.707 #
17)	L4 AR-1242-2	5.536	4.697	151223	95853	1.985	1.995
18)	L4 AR-1242-3	5.590	0.000	237425	0	4.846	N.D. #
19)	L4 AR-1242-4	5.686	0.000	73242	0	1.908	N.D. #
20)	L4 AR-1242-5	6.428	5.495	67559	135801	1.656	4.449 #
21)	L5 AR-1248-1	5.509	4.697	110372	95853	2.619	3.369 #
22)	L5 AR-1248-2	5.787	0.000	68076	0	1.077	N.D. #
23)	L5 AR-1248-3	5.982	0.000	135697	0	1.937	N.D. #
24)	L5 AR-1248-4	6.414f	5.127	234066	259784	3.319	5.399 #
25)	L5 AR-1248-5	6.428	0.000	67559	0	0.959	N.D. #
26)	L6 AR-1254-1	6.377	5.495	126642	135801	1.634	2.074 #
27)	L6 AR-1254-2	6.590	5.653	300414	213230	2.582	3.656 #
28)	L6 AR-1254-3	6.961	6.039	477957	258578	4.081	2.939 #
29)	L6 AR-1254-4	7.250	6.270	18422	207613	0.232	4.532 #
30)	L6 AR-1254-5	7.657	6.703	56636	39333	0.581	0.505
31)	L7 AR-1260-1	7.128	6.182	168020	333386	1.740	5.111 #
32)	L7 AR-1260-2	7.403f	6.381	309187	158128	2.799	2.100
33)	L7 AR-1260-3	7.743	6.528	30412	563393	0.402	7.773 #
34)	L7 AR-1260-4	7.955	6.992	723600	47578	7.710	0.877 #
35)	L7 AR-1260-5	8.282	7.237	135912	121965	0.857	1.020
36)	L8 AR-1262-1	7.743	6.787	30412	96022	0.254	2.692 #
37)	L8 AR-1262-2	8.282	6.992	135912	47578	0.709	0.612
38)	L8 AR-1262-3	8.600	7.514	207052	292436	1.494	4.642 #
39)	L8 AR-1262-4	8.689	7.586	581511	463319	9.344	4.052 #
40)	L8 AR-1262-5	9.366f	8.094	31194	99798	0.441	2.003 #
41)	L9 AR-1268-1	8.600	7.514	207052	292436	0.848	1.616 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121522\
 Data File : PP053811.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Dec 2022 23:46
 Operator : YP\AJ
 Sample : N6037-21
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 16 01:53:08 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121322.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Dec 14 05:38:40 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.689	7.586	581511	463319	2.559	2.715
43) L9	AR-1268-3	8.929	0.000	97962	0	0.511	N.D. #
44) L9	AR-1268-4	9.366f	8.094	31194	99798	0.405	1.791 #
45) L9	AR-1268-5	9.764	8.374	173525	326628	0.295	0.691 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121522\
Data File : PP053811.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Dec 2022 23:46
Operator : YP\AJ
Sample : N6037-21
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 16 01:53:08 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121322.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Dec 14 05:38:40 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.340 min
Delta R.T.: 0.000 min
Response: 43479618
Conc: 20.71 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.592 min
Delta R.T.: -0.001 min
Response: 31093977
Conc: 19.91 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.107 min
Delta R.T.: 0.005 min
Response: 23599231
Conc: 13.90 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.633 min
Delta R.T.: -0.003 min
Response: 19703692
Conc: 14.11 ng/ml



#3 AR-1016-1


R.T.: 5.509 min
Delta R.T.: -0.003 min
Response: 110372
Conc: 1.63 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.697 min
Delta R.T.: 0.011 min
Response: 95853
Conc: 2.09 ng/ml




#4 AR-1016-2

R.T.: 5.536 min
Delta R.T.: 0.001 min
Response: 151223
Conc: 1.52 ng/ml



#4 AR-1016-2


R.T.: 4.697 min
Delta R.T.: -0.008 min
Response: 95853
Conc: 1.53 ng/ml



#5 AR-1016-3


R.T.: 5.590 min
Delta R.T.: -0.006 min
Response: 237425
Conc: 3.76 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 0.000 min
Exp R.T. : 4.882 min
Response: 0
Conc: N.D.




#6 AR-1016-4

R.T.: 5.686 min
Delta R.T.: -0.009 min
Response: 73242
Conc: 1.48 ng/ml



#6 AR-1016-4


R.T.: 0.000 min
Exp R.T. : 4.924 min
Response: 0
Conc: N.D.



#7 AR-1016-5


R.T.: 5.982 min
Delta R.T.: -0.009 min
Response: 135697
Conc: 2.58 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.127 min
Delta R.T.: -0.012 min
Response: 259784
Conc: 7.16 ng/ml




#8 AR-1221-1

R.T.: 4.547 min
Delta R.T.: -0.002 min
Response: 1299396
Conc: 50.44 ng/ml



#8 AR-1221-1


R.T.: 3.815 min
Delta R.T.: 0.005 min
Response: 147570
Conc: 7.62 ng/ml



#9 AR-1221-2


R.T.: 4.647 min
Delta R.T.: 0.012 min
Response: 649121
Conc: 33.34 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.892 min
Delta R.T.: -0.003 min
Response: 1008964
Conc: 69.02 ng/ml




#10 AR-1221-3

R.T.: 4.701 min
Delta R.T.: -0.011 min
Response: 38307
Conc: 0.66 ng/ml



#10 AR-1221-3


R.T.: 0.000 min
Exp R.T. : 3.972 min
Response: 0
Conc: N.D.



#11 AR-1232-1


R.T.: 4.701 min
Delta R.T.: -0.010 min
Response: 38307
Conc: 0.86 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 0.000 min
Exp R.T. : 3.971 min
Response: 0
Conc: N.D.




#12 AR-1232-2

R.T.: 5.251 min
Delta R.T.: 0.009 min
Response: 77333
Conc: 3.01 ng/ml



#12 AR-1232-2


R.T.: 4.697 min
Delta R.T.: -0.008 min
Response: 95853
Conc: 3.21 ng/ml



#13 AR-1232-3


R.T.: 5.536 min
Delta R.T.: 0.002 min
Response: 151223
Conc: 3.22 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 0.000 min
Exp R.T. : 4.882 min
Response: 0
Conc: N.D.




#14 AR-1232-4

R.T.: 5.686 min
Delta R.T.: -0.009 min
Response: 73242
Conc: 3.09 ng/ml



#14 AR-1232-4


R.T.: 0.000 min
Exp R.T. : 4.966 min
Response: 0
Conc: N.D.



#15 AR-1232-5


R.T.: 5.787 min
Delta R.T.: 0.000 min
Response: 68076
Conc: 3.69 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.127 min
Delta R.T.: -0.011 min
Response: 259784
Conc: 15.78 ng/ml




#16 AR-1242-1

R.T.: 5.509 min
Delta R.T.: -0.003 min
Response: 110372
Conc: 2.11 ng/ml



#16 AR-1242-1


R.T.: 4.697 min
Delta R.T.: 0.011 min
Response: 95853
Conc: 2.71 ng/ml



#17 AR-1242-2


R.T.: 5.536 min
Delta R.T.: 0.001 min
Response: 151223
Conc: 1.98 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.697 min
Delta R.T.: -0.008 min
Response: 95853
Conc: 2.00 ng/ml




#18 AR-1242-3

R.T.: 5.590 min
Delta R.T.: -0.007 min
Response: 237425
Conc: 4.85 ng/ml



#18 AR-1242-3


R.T.: 0.000 min
Exp R.T. : 4.882 min
Response: 0
Conc: N.D.



#19 AR-1242-4


R.T.: 5.686 min
Delta R.T.: -0.009 min
Response: 73242
Conc: 1.91 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 0.000 min
Exp R.T. : 4.966 min
Response: 0
Conc: N.D.




#20 AR-1242-5

R.T.: 6.428 min
Delta R.T.: -0.007 min
Response: 67559
Conc: 1.66 ng/ml



#20 AR-1242-5


R.T.: 5.495 min
Delta R.T.: 0.003 min
Response: 135801
Conc: 4.45 ng/ml



#21 AR-1248-1


R.T.: 5.509 min
Delta R.T.: -0.003 min
Response: 110372
Conc: 2.62 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.697 min
Delta R.T.: 0.011 min
Response: 95853
Conc: 3.37 ng/ml




#22 AR-1248-2

R.T.: 5.787 min
Delta R.T.: 0.000 min
Response: 68076
Conc: 1.08 ng/ml



#22 AR-1248-2


R.T.: 0.000 min
Exp R.T.: 4.924 min
Response: 0
Conc: N.D.



#23 AR-1248-3


R.T.: 5.982 min
Delta R.T.: -0.009 min
Response: 135697
Conc: 1.94 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 0.000 min
Exp R.T. : 4.966 min
Response: 0
Conc: N.D.




#24 AR-1248-4

R.T.: 6.414 min
Delta R.T.: 0.017 min
Response: 234066
Conc: 3.32 ng/ml



#24 AR-1248-4


R.T.: 5.127 min
Delta R.T.: -0.012 min
Response: 259784
Conc: 5.40 ng/ml



#25 AR-1248-5


R.T.: 6.428 min
Delta R.T.: -0.007 min
Response: 67559
Conc: 0.96 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 0.000 min
Exp R.T. : 5.533 min
Response: 0
Conc: N.D.




#26 AR-1254-1

R.T.: 6.377 min
Delta R.T.: 0.008 min
Response: 126642
Conc: 1.63 ng/ml



#26 AR-1254-1


R.T.: 5.495 min
Delta R.T.: 0.003 min
Response: 135801
Conc: 2.07 ng/ml



#27 AR-1254-2


R.T.: 6.590 min
Delta R.T.: 0.000 min
Response: 300414
Conc: 2.58 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.653 min
Delta R.T.: 0.012 min
Response: 213230
Conc: 3.66 ng/ml




#28 AR-1254-3

R.T.: 6.961 min
Delta R.T.: 0.004 min
Response: 477957
Conc: 4.08 ng/ml



#28 AR-1254-3


R.T.: 6.039 min
Delta R.T.: -0.008 min
Response: 258578
Conc: 2.94 ng/ml



#29 AR-1254-4


R.T.: 7.250 min
Delta R.T.: 0.006 min
Response: 18422
Conc: 0.23 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.270 min
Delta R.T.: -0.006 min
Response: 207613
Conc: 4.53 ng/ml




#30 AR-1254-5

R.T.: 7.657 min
Delta R.T.: -0.006 min
Response: 56636
Conc: 0.58 ng/ml



#30 AR-1254-5


R.T.: 6.703 min
Delta R.T.: 0.007 min
Response: 39333
Conc: 0.50 ng/ml



#31 AR-1260-1


R.T.: 7.128 min
Delta R.T.: 0.005 min
Response: 168020
Conc: 1.74 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.182 min
Delta R.T.: 0.004 min
Response: 333386
Conc: 5.11 ng/ml




#32 AR-1260-2

R.T.: 7.403 min
Delta R.T.: 0.021 min
Response: 309187
Conc: 2.80 ng/ml



#32 AR-1260-2


R.T.: 6.381 min
Delta R.T.: 0.014 min
Response: 158128
Conc: 2.10 ng/ml



#33 AR-1260-3


R.T.: 7.743 min
Delta R.T.: 0.000 min
Response: 30412
Conc: 0.40 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.528 min
Delta R.T.: 0.006 min
Response: 563393
Conc: 7.77 ng/ml




#34 AR-1260-4

R.T.: 7.955 min
Delta R.T.: -0.013 min
Response: 723600
Conc: 7.71 ng/ml



#34 AR-1260-4


R.T.: 6.992 min
Delta R.T.: -0.005 min
Response: 47578
Conc: 0.88 ng/ml



#35 AR-1260-5


R.T.: 8.282 min
Delta R.T.: -0.003 min
Response: 135912
Conc: 0.86 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.237 min
Delta R.T.: -0.002 min
Response: 121965
Conc: 1.02 ng/ml




#36 AR-1262-1

R.T.: 7.743 min
Delta R.T.: 0.001 min
Response: 30412
Conc: 0.25 ng/ml



#36 AR-1262-1


R.T.: 6.787 min
Delta R.T.: -0.001 min
Response: 96022
Conc: 2.69 ng/ml



#37 AR-1262-2


R.T.: 8.282 min
Delta R.T.: -0.003 min
Response: 135912
Conc: 0.71 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 6.992 min
Delta R.T.: -0.005 min
Response: 47578
Conc: 0.61 ng/ml




#38 AR-1262-3

R.T.: 8.600 min
Delta R.T.: 0.002 min
Response: 207052
Conc: 1.49 ng/ml



#38 AR-1262-3


R.T.: 7.514 min
Delta R.T.: -0.011 min
Response: 292436
Conc: 4.64 ng/ml



#39 AR-1262-4


R.T.: 8.689 min
Delta R.T.: 0.002 min
Response: 581511
Conc: 9.34 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.586 min
Delta R.T.: -0.002 min
Response: 463319
Conc: 4.05 ng/ml




#40 AR-1262-5

R.T.: 9.366 min
Delta R.T.: 0.021 min
Response: 31194
Conc: 0.44 ng/ml



#40 AR-1262-5


R.T.: 8.094 min
Delta R.T.: 0.010 min
Response: 99798
Conc: 2.00 ng/ml



#41 AR-1268-1


R.T.: 8.600 min
Delta R.T.: 0.005 min
Response: 207052
Conc: 0.85 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.514 min
Delta R.T.: -0.012 min
Response: 292436
Conc: 1.62 ng/ml




#42 AR-1268-2

R.T.: 8.689 min
Delta R.T.: 0.000 min
Response: 581511
Conc: 2.56 ng/ml



#42 AR-1268-2


R.T.: 7.586 min
Delta R.T.: -0.003 min
Response: 463319
Conc: 2.72 ng/ml



#43 AR-1268-3


R.T.: 8.929 min
Delta R.T.: 0.011 min
Response: 97962
Conc: 0.51 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 0.000 min
Exp R.T. : 7.799 min
Response: 0
Conc: N.D.




#44 AR-1268-4

R.T.: 9.366 min
Delta R.T.: 0.022 min
Response: 31194
Conc: 0.41 ng/ml



#44 AR-1268-4


R.T.: 8.094 min
Delta R.T.: 0.009 min
Response: 99798
Conc: 1.79 ng/ml



#45 AR-1268-5

R.T.: 9.764 min
Delta R.T.: 0.003 min
Response: 173525
Conc: 0.30 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.374 min
Delta R.T.: -0.004 min
Response: 326628
Conc: 0.69 ng/ml