

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121620\
 Data File : PP032177.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 16 Dec 2020 9:04
 Operator : DD\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 17 00:59:02 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121120.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Dec 12 05:15:55 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.778	4.048	2382326	2141358	46.247	45.275
2) SA Decachlor...	10.637	9.338	1591047	1733745	50.736	51.698
Target Compounds						
3) L1 AR-1016-1	6.081	5.298	698954	614234	479.208	469.904
4) L1 AR-1016-2	6.104	5.319	1056492	926736	470.342	468.076
5) L1 AR-1016-3	6.170	5.510	646229	508555	467.550	475.572
6) L1 AR-1016-4	6.275	5.561	543818	390275	471.586	459.951
7) L1 AR-1016-5	6.589	5.789	504379	482760	472.713	444.278
31) L7 AR-1260-1	7.759	6.883	808903	888859	466.739	468.844
32) L7 AR-1260-2	8.025	7.079	1023986	1062734	498.389	488.529
33) L7 AR-1260-3	8.390	7.234	841111	1023469	482.023	481.707
34) L7 AR-1260-4	8.618	7.715	872904	871344	484.407	490.455
35) L7 AR-1260-5	8.932	7.963	1867898	2047808	519.596	535.307



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121620\
Data File : PP032177.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 16 Dec 2020 9:04
Operator : DD\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 17 00:59:02 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121120.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Dec 12 05:15:55 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.778 min
Delta R.T.: -0.004 min
Response: 2382326
Conc: 46.25 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.048 min
Delta R.T.: -0.006 min
Response: 2141358
Conc: 45.28 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.637 min
Delta R.T.: -0.010 min
Response: 1591047
Conc: 50.74 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.338 min
Delta R.T.: -0.010 min
Response: 1733745
Conc: 51.70 ng/ml



#3 AR-1016-1


R.T.: 6.081 min
Delta R.T.: -0.004 min
Response: 698954
Conc: 479.21 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 5.298 min
Delta R.T.: -0.008 min
Response: 614234
Conc: 469.90 ng/ml




#4 AR-1016-2

R.T.: 6.104 min
Delta R.T.: -0.003 min
Response: 1056492
Conc: 470.34 ng/ml



#4 AR-1016-2


R.T.: 5.319 min
Delta R.T.: -0.007 min
Response: 926736
Conc: 468.08 ng/ml



#5 AR-1016-3


R.T.: 6.170 min
Delta R.T.: -0.004 min
Response: 646229
Conc: 467.55 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 5.510 min
Delta R.T.: -0.008 min
Response: 508555
Conc: 475.57 ng/ml




#6 AR-1016-4

R.T.: 6.275 min
Delta R.T.: -0.004 min
Response: 543818
Conc: 471.59 ng/ml



#6 AR-1016-4


R.T.: 5.561 min
Delta R.T.: -0.008 min
Response: 390275
Conc: 459.95 ng/ml



#7 AR-1016-5


R.T.: 6.589 min
Delta R.T.: -0.004 min
Response: 504379
Conc: 472.71 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.789 min
Delta R.T.: -0.009 min
Response: 482760
Conc: 444.28 ng/ml




#31 AR-1260-1

R.T.: 7.759 min
Delta R.T.: -0.005 min
Response: 808903
Conc: 466.74 ng/ml



#31 AR-1260-1


R.T.: 6.883 min
Delta R.T.: -0.008 min
Response: 888859
Conc: 468.84 ng/ml



#32 AR-1260-2


R.T.: 8.025 min
Delta R.T.: -0.005 min
Response: 1023986
Conc: 498.39 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 7.079 min
Delta R.T.: -0.008 min
Response: 1062734
Conc: 488.53 ng/ml




#33 AR-1260-3

R.T.: 8.390 min
Delta R.T.: -0.005 min
Response: 841111
Conc: 482.02 ng/ml



#33 AR-1260-3


R.T.: 7.234 min
Delta R.T.: -0.009 min
Response: 1023469
Conc: 481.71 ng/ml



#34 AR-1260-4


R.T.: 8.618 min
Delta R.T.: -0.005 min
Response: 872904
Conc: 484.41 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.715 min
Delta R.T.: -0.009 min
Response: 871344
Conc: 490.45 ng/ml



#35 AR-1260-5

R.T.: 8.932 min
Delta R.T.: -0.005 min
Response: 1867898
Conc: 519.60 ng/ml



#35 AR-1260-5

R.T.: 7.963 min
Delta R.T.: -0.009 min
Response: 2047808
Conc: 535.31 ng/ml