

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121620\
 Data File : PP032190.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 16 Dec 2020 14:26
 Operator : DD\AJ
 Sample : L5113-02
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 SS1

Manual Integrations
 APPROVED

Ankita
 12/17/2020 12:42:13 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 17 01:06:33 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121120.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Dec 12 05:15:55 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.777	4.049	668244	590090	12.972m	12.476
2) SA Decachlor...	10.635	9.338	346942	391691	11.063	11.680
Target Compounds						
26) L6 AR-1254-1	6.983	6.169	119018	165464	76.626	80.049
27) L6 AR-1254-2	7.214	6.327	185487	135125	80.636	75.374
28) L6 AR-1254-3	7.591	6.747	176677	216879	72.559m	75.885m
29) L6 AR-1254-4	7.885	6.986	123103	76658	79.978	55.471m#
30) L6 AR-1254-5	8.311	7.414	188191	703859	109.805m	295.862 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121620\
Data File : PP032190.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 16 Dec 2020 14:26
Operator : DD\AJ
Sample : L5113-02
Misc :
ALS Vial : 10 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampled :
SS1


Manual Integrations
APPROVED

Ankita
12/17/2020 12:42:13 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 17 01:06:33 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121120.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Dec 12 05:15:55 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.777 min
Delta R.T.: -0.005 min
Response: 668244
Conc: 12.97 ng/ml m

Instrument :
ECD_P
ClientSampleId :
SS1


Manual Integrations
APPROVED

Ankita
12/17/2020 12:42:13 PM




#1 Tetrachloro-m-xylene

R.T.: 4.049 min
Delta R.T.: -0.005 min
Response: 590090
Conc: 12.48 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.635 min
Delta R.T.: -0.011 min
Response: 346942
Conc: 11.06 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.338 min
Delta R.T.: -0.010 min
Response: 391691
Conc: 11.68 ng/ml




#26 AR-1254-1

R.T.: 6.983 min
Delta R.T.: -0.004 min
Response: 119018
Conc: 76.63 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS1


Manual Integrations
APPROVED

Ankita
12/17/2020 12:42:13 PM




#26 AR-1254-1

R.T.: 6.169 min
Delta R.T.: -0.005 min
Response: 165464
Conc: 80.05 ng/ml




#27 AR-1254-2

R.T.: 7.214 min
Delta R.T.: -0.003 min
Response: 185487
Conc: 80.64 ng/ml



#27 AR-1254-2

R.T.: 6.327 min
Delta R.T.: -0.006 min
Response: 135125
Conc: 75.37 ng/ml




#28 AR-1254-3

R.T.: 7.591 min
Delta R.T.: -0.004 min
Response: 176677
Conc: 72.56 ng/ml m

Instrument :
ECD_P
ClientSampleId :
SS1


Manual Integrations
APPROVED

Ankita
12/17/2020 12:42:13 PM




#28 AR-1254-3

R.T.: 6.747 min
Delta R.T.: -0.006 min
Response: 216879
Conc: 75.88 ng/ml m




#29 AR-1254-4

R.T.: 7.885 min
Delta R.T.: -0.006 min
Response: 123103
Conc: 79.98 ng/ml



#29 AR-1254-4

R.T.: 6.986 min
Delta R.T.: -0.007 min
Response: 76658
Conc: 55.47 ng/ml m




#30 AR-1254-5

R.T.: 8.311 min
Delta R.T.: -0.007 min
Response: 188191
Conc: 109.81 ng/ml m

Instrument :
ECD_P
ClientSampleId :
SS1

Manual Integrations
APPROVED

Ankita
12/17/2020 12:42:13 PM



#30 AR-1254-5

R.T.: 7.414 min
Delta R.T.: -0.007 min
Response: 703859
Conc: 295.86 ng/ml